American National Standard

ANSI/AAMI CI86:2017/(R)2025

Cochlear implant systems: Requirements for safety, functional verification, labeling and reliability reporting

AAMI
Advancing Safety in Health Technology

Cochlear implant systems: Requirements for safety, functional verification, labeling and reliability reporting

Approved 15 December 2016 by
AAMI

Approved 6 January 2017 and reaffirmed 5 May 2025 by
American National Standards Institute, Inc.

Abstract: This standard establishes minimum requirements for those active implantable medical devices known as cochlear implants or cochlear prostheses, which are intended to treat hearing impairment by means of electrical stimulation of the cochlea. Devices that treat hearing impairment other than by including electrical stimulation of the cochlea are not covered by this standard. This standard applies to the electrical stimulation component(s) of combination devices that treat hearing impairment using multiple means, including electrical stimulation. The tests specified in this standard are industry-accepted tests and are to be carried out on samples of devices to show compliance. This standard is also applicable to non-implantable parts and accessories of the devices, including fitting and diagnostic components. General and specific requirements are provided with regard to design verification, post-implantation device testing, reliability assessment and reporting, packaging and labeling, protections of the patient associated with design issues and device malfunctions, and protections of the device associated with environmental challenges arising from transport, storage, handling during implantation, unrelated-medical treatments, and normal use.

Keywords: cochlear implants, cochlear prosthesis, medical device

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than 5 years from the date of publication. Interested parties may obtain current information on all AAMI standards and recommended practices by calling or writing AAMI.

All AAMI standards, recommended practices, technical information reports, and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

The transition period for voluntary compliance with this standard is indicated below.

- Design and test requirements (transition period is 2 years):
  - For devices or changes to devices approved prior to publication of standard, conformance to this standard is optional.
  - For devices or changes approved during the transition period, conformance will be negotiated between each manufacturer and the regulatory agency.
  - For new devices or designs marketed after the transition period, conformance to this standard is mandatory if a cochlear manufacturer intends to claim compliance to this standard.
- Reliability reporting:
  - For all implantable components (both currently marketed and no longer distributed devices), the transition period for reporting as specified in this standard is 2 years from publication.
  - For implantable components that will be introduced after publication, the transition period for reporting as specified in this standard is 1 year.
  - For all sound processors (reporting as specified in Section 11), the transition period for reporting as specified in this standard is 2 years from publication.

NOTE—This document contains active links. To return to your previous location after following an active link, use Alt (left arrow).

Published by

AAMI
901 N. Glebe Road, Suite 300
Arlington, VA 22203

© 2017 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of $100,000 per offense. For permission regarding the use of all or any part of this document, contact the Copyright Clearance Center.

Printed in the United States of America

ISBN 978-1-57020-662-7

Contents

Page

Glossary of equivalent standards v
Committee representation vi
Acknowledgments vi
Foreword vii
1 Scope 1
2 Normative references 2
3 Definitions 4
4 Units, abbreviations, and symbols 12
5 General requirements for characterizing a cochlear implant system 14
6 General requirements for implantable parts 27
7 General requirements for non-implantable parts 29
8 System-level inspection, characterization, and measurement 31
9 Implantation support (RESERVED FOR FUTURE VERSIONS) 42
10 Post-implantation testing, *in vivo* assessment, and analysis of failed devices 42
11 Reliability monitoring and reporting 45
12 Information on use, warnings, and hazards 64
13 General arrangement of the packaging system 70
14 Markings on the packaging system and the cochlear implant device 70
15 Safety from unintentional biological effects of the cochlear implant system 76
16 Safety of secondary features of the cochlear implant system 76
17 Safety of electrical stimulation 77
18 Safety of implantable energy sources 82
19 Safety from heat sources 83
20 Safety from unintended effects caused by the device 85
21 Safety and device immunity during magnetic resonance imaging 87
22 Device immunity to external stresses caused by medical treatment and procedures 92
23 Device immunity to stresses of mechanical forces 97
24 Device immunity to stresses caused by atmospheric pressure changes 107
25 Device immunity to stresses caused by temperature changes 109
26 Non-implantable device immunity to stresses caused by environment and usage 110
Annex A (informative) Clinical identification and management of cochlear implant device failures 123
Annex B (informative) Clinical checklist prior to explantation Signs and symptoms checklist 125
Annex C (informative) Returned implant analysis 129
Annex D (informative) Indications of performance decline 135
Annex E (informative) Reliability reporting to regulatory bodies 136
Annex F (informative) Pareto analysis 145
Annex G (informative) Failed component return rate (FCRR) graphic and table 147

Annex H (informative) Reliability reporting template for the public and clinical community.....149

Annex I (informative) Product specification data sheets 152

Annex J (informative) Mechanical testing of leads and interfaces to case bodies.....158

Annex K (informative) Logic flow diagram of the relationships between the required testing, explant category classification, and reliability reporting to both regulatory bodies and the public 161

Bibliography.....167

Glossary of equivalent standards

International Standards adopted in the United States may include normative references to other International Standards. AAMI maintains a current list of each International Standard that has been adopted by AAMI (and ANSI). Available on the AAMI website through the link below, this glossary listing gives the corresponding U.S. designation and level of equivalency to the International Standard.

www.aami.org/standards/glossary.pdf

Committee representation

Association for the Advancement of Medical Instrumentation
Cochlear Implant Committee

This standard was developed by the AAMI Cochlear Implant Committee. Committee approval of the standard does not necessarily mean that all committee members voted for its approval.

At the time this standard was published, the **AAMI Cochlear Implant Committee** had the following members:

Co-Chairs:
Cedric Mark Navarro
Julie Verhoff, AuD, PhD

Members:
Douglas D. Backous, MD, Swedish Medical Center
Kevin Crowe, BSc(Eng), Cochlear Ltd.
Vasant Dasika, PhD, FDA/CDRH
Bruce J. Gantz, MD, University of Iowa Hospitals and Clinics
Eric Khosravi, MAS, MSQA, Oticon Medical
Bomjun Kwon, PhD, Gallaudet University
Anil K. Lalwani, MD, American Neurotology Society
Cedric Mark Navarro, Advanced Bionics LLC
Michael Pecht, PhD, University of Maryland College Park
Mark J. Syms, MD FACS, Arizona Ear Center
Julie Verhoff, AuD, PhD, The River School / Chattering Children
Fan-Gang Zeng, PhD, University of California at Irvine
Martin Zimmerling, PhD, MED-EL GmbH
Teresa A. Zwolan, PhD, University of Michigan

Alternates:
Sean Bundy, Cochlear Ltd
Charles Finley, PhD, Advanced Bionics LLC
Srinivas Nandkumar, PhD, FDA/CDRH
Edward Overstreet, PhD, Oticon Medical
Manfred Pieber, MSc, MED-EL GmbH
William F. Regnault, PhD, FDA/CDRH
Bhanu Sood, MSc
Liliane L. Tessa, BS, University of Maryland

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

Acknowledgments

The committee gratefully acknowledges James (Jim) Kane, PhD, and William (Bill) Regnault, PhD, of the U.S. Food and Drug Administration for their contributions to the development of an American National Standard for cochlear implants. Drawing on their experience from their regulatory careers, they have broadened the standard to better promote safe and efficacious applications of cochlear implants. Jim served as chief advocate and co-chairman of the committee from its inception to his retirement in 2013. Bill's extensive technical expertise and critical thinking influenced numerous aspects of this standard. Their commitment of time and effort attests to their dedication to patient welfare and is greatly appreciated.

Foreword

This standard was developed by the AAMI Cochlear Implant Committee and establishes minimum safety and performance requirements for cochlear implants based on industry-accepted test methods. It is intended to assist manufacturers in testing and labeling their devices and in reporting device system reliability to the professional health care community and public. In turn, such uniformity and consistency should facilitate more informed decision making by health care professionals, potential cochlear implant patients, and parents of potential pediatric cochlear implant recipients, when selecting applicable device systems.

Cochlear implant systems consist of an implantable device that electrically stimulates the inner ear to produce perceptions of sound in patients with significantly impaired acoustic hearing and external devices that convert acoustic sound signals to electrical signals, communicate electromagnetically with the implant, and program the device. In the United States, cochlear implant systems are medical devices and, under the Food, Drug, and Cosmetic Act, as amended by the Medical Device Amendments of May 28, 1976, are subject to regulation by the U.S. Food and Drug Administration (FDA). Such regulation includes, but is not limited to, FDA requirements for premarket approval (Section 515 of the Act) and medical device reporting.

As used within the context of this document, “shall” indicates requirements strictly to be followed to conform to the standard. “Should” indicates that among several possibilities, one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action should be avoided but is not prohibited. “May” is used to indicate that a course of action is permissible within the limits of the standard. “Can” is used as a statement of possibility and capability.

AAMI and ANSI procedures require that standards be reviewed every five years and, if necessary, revised to reflect technological advances that may have occurred since publication.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments and knowledge gained from field experience with the devices. To remain relevant, it must be modified as technological advances are made and as new data come to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 901 N. Glebe Road, Suite 300, Arlington, VA 22203.

NOTE—This foreword does not contain provisions of the American National Standard, *Cochlear implant systems: Requirements for safety, functional verification, labeling and reliability reporting* (ANSI/AAMI CI86:2017), but it does provide important information about the development and intended use of the document.

Cochlear implant systems: Requirements for safety, functional verification, labeling and reliability reporting

1 Scope

This standard specifies requirements, test procedures, methods and labeling for active implantable medical devices intended to treat hearing impairment by means of electrical stimulation of the cochlea. Such devices are referred to as cochlear implants or cochlear prostheses. This standard is also applicable to non-implantable parts and accessories of the devices, including fitting and diagnostic components.

A cochlear implant system can be a single device, a combination of devices, or a combination of a device or devices and one or more accessories. Not all of these parts are required to be either partially or totally implantable, but some requirements for non-implantable parts and accessories must be specified if they could affect the safety or performance of the implantable part.

Devices that treat hearing impairment exclusively by means other than direct electrical stimulation are not covered by this standard (e.g., hearing aids, bone conduction devices, middle ear prostheses). For devices that treat hearing impairment using combined electrical and acoustic stimulation, only the components relevant to electrical stimulation are covered by this standard. The acoustic stimulation is covered by other relevant standards. For devices that deliver medicinal substances and are consequently combination products for regulatory purposes, this standard covers only the components relevant to electrical stimulation.

The tests specified in this standard serve two purposes:

a) To demonstrate compliance with specific requirements in the standard (type tests) and
b) To demonstrate functional performance of the system and/or subsystem components at specified reliability levels within known confidence limits (reliability tests)

The sample sizes and selection methods vary across the tests, depending on the purpose of the test, the degree of patient risk associated with failure of the component being tested, and the desired degree of confidence for the test.

The electrical, mechanical, and thermal characteristics of the implantable part shall be determined either by the appropriate method detailed in this particular standard or by another method demonstrated to have accuracy equal to, or better than, the method specified. In the case of dispute, the method detailed in this particular standard shall apply.

This standard requires the reporting of information regarding the design, operation, testing, and reliability of the implantable and non-implantable components of the cochlear implant system, including accessories required for or which extend the operation of the system. All such information shall be considered confidential to the manufacturer unless the standard states specifically that the information is for broader distribution and the standard provides detailed requirements for that broader distribution.

Throughout this standard, descriptive or explanatory information is included with the text describing requirements and tests. The information contained in notes should be considered in the implementation of tests and/or in the interpretation of results.

This standard does not address all of the appropriate aspects of wireless technology incorporated into cochlear implant systems. The standard does include some tests for specific aspects of wireless functionality of the system and its components.

This standard does not address the clinical/audiological performance data that a manufacturer might include as part of a regulatory submission to support the safety and effectiveness of a device.

© 2017 Association for the Advancement of Medical Instrumentation ■ AAMI CI86:2017