AS ISO 14155.1—2004
ISO 14155-1:2003

AS ISO 14155.1—2004

Australian Standard™

Clinical investigations of medical devices for human subjects

Part 1: General requirements

STANDARDS AUSTRALIA

This Australian Standard was prepared by Committee HE-012, Surgical Implants. It was approved on behalf of the Council of Standards Australia on 17 June 2004.
This Standard was published on 27 July 2004.

The following are represented on Committee HE-012:

Australian College of Operating Room Nurses
Australian Dental Association
Australian Industry Group
Australian Orthopaedic Association
Commonwealth Department of Health and Ageing
Department of Defence (Australia)
Medical Industry Association of Australia Inc
Neurosurgical Society of Australasia
Royal Australasian College of Surgeons
Royal Perth Hospital
The Australian Society for Biomaterials
University of New South Wales
University of Sydney

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Web Shop at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

Australian Standards™ and other products and services developed by Standards Australia are published and distributed under contract by SAI Global, which operates the Standards Web Shop.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 03517.

Australian Standard™

Clinical investigations of medical devices for human subjects

Part 1: General requirements

Originated as AS ISO 14155—2002.
Revised and redesignated as AS ISO 14155.1—2004.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 6169 0

PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee HE-012, Surgical Implants, to supersede AS ISO 14155—2002, *Clinical investigation of medical devices.* After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard.

This Standard is identical with, and has been reproduced from, ISO 14155-1:2003, *Clinical investigation of medical devices for human subjects,* Part 1: *General requirements.*

The objective of this Standard is to specify requirements for the conduct of a clinical investigation to establish the performance of medical devices intended to mimic normal clinical use, reveal adverse events under normal conditions of use, and permit assessment of the acceptable risks.

The term ‘informative’ has been used in this Standard to define the application of the annex to which it applies. An informative annex is only for information and guidance.

As this Standard is reproduced from an international Standard, the following applies:

(a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.

(b) In the source text ‘this part of ISO 14155’ should read ‘this Australian Standard’.

(c) A full point substitutes for a comma when referring to a decimal marker.

(d) The normative reference in Clause 2 has not been adopted as an Australian Standard.

CONTENTS

Introduction v

1 Scope 1

2 Normative references 1

3 Terms and definitions 2

4 Justification for a clinical investigation 5

5 Ethical considerations 5
5.1 Declaration of Helsinki 5
5.2 Improper influence or inducement 5
5.3 Compensation and additional health care 5
5.4 Responsibilities 5

6 General requirements 5
6.1 Formal agreement(s) 5
6.2 Qualifications 5
6.3 Clinical investigation plan 6
6.4 Design of the clinical investigation 6
6.5 Confidentiality 6
6.6 Start of clinical investigation 6
6.7 Informed consent 6
6.8 Suspension or early termination of the clinical investigation 8
6.9 Document and data control 8
6.10 Accounting for subjects 9
6.11 Access to preclinical and clinical information 9
6.12 Auditing 9

7 Documentation 9
7.1 General 9
7.2 Clinical investigator's brochure 9
7.3 Other documents 10

8 Sponsor 10
8.1 General 10
8.2 Responsibilities of sponsor 10

9 Monitor 11
9.1 Responsibilities of monitor 11

10 Clinical investigator 12
10.1 General 12
10.2 Qualification of clinical investigator 12
10.3 Responsibilities of clinical investigator 12

11 Final report 14
11.1 Presentation of results 14
11.2 Contents of the final report 14

Annex A (informative) **Suggested procedure for literature review** **15**
Annex B (informative) **Information for the ethics committees** **17**
Annex C (informative) **Final reports of clinical investigations with medical devices** **18**
Bibliography **21**

INTRODUCTION

This part of ISO 14155 is intended to be applied worldwide to clinical investigations of medical devices in order to fulfil the technical aspects of the various national, regional and international regulatory requirements. As the legal regulatory requirements presently differ throughout the world, regulatory specifics have been excluded from the scope of this part of ISO 14155. They are part of national or regional legislative texts and can be referenced in the national or regional forewords, as appropriate.

NOTES

AUSTRALIAN STANDARD

Clinical investigation of medical devices for human subjects —

Part 1:
General requirements

1 Scope

This part of ISO 14155 defines procedures for the conduct and performance of clinical investigations of medical devices. It specifies general requirements intended to

— protect human subjects,

— ensure the scientific conduct of the clinical investigation,

— assist sponsors, monitors, investigators, ethics committees, regulatory authorities and bodies involved in the conformity assessment of medical devices.

This part of ISO 14155

a) specifies requirements for the conduct of a clinical investigation such that it establishes the performance of the medical device during the clinical investigation intended to mimic normal clinical use, reveals adverse events under normal conditions of use, and permits assessment of the acceptable risks having regard to the intended performance of the medical device,

b) specifies requirements for the organization, conduct, monitoring, data collection and documentation of the clinical investigation of a medical device,

c) is applicable to all clinical investigation(s) of medical devices whose clinical performance and safety is being assessed in human subjects.

This part of ISO 14155 is not applicable to *in vitro* diagnostic medical devices.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14155-2, *Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans*

© Standards Australia