PD ISO/TR 6057:2023



BSI Standards Publication

Hydraulic fluid power — Sample calculations for ISO 11171

bsi.

National foreword

This Published Document is the UK implementation of ISO/TR 6057:2023.

The UK participation in its preparation was entrusted to Technical Committee MCE/18/-/6, Contamination control.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication is not to be regarded as a British Standard.

© The British Standards Institution 2023
Published by BSI Standards Limited 2023

ISBN 978 0 539 21776 6

ICS 23.100.60

Compliance with a Published Document cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 April 2023.

Amendments/corrigenda issued since publication

Date	Text affected

PD ISO/TR 6057:2023

TECHNICAL REPORT

ISO/TR 6057

First edition
2023-04

Hydraulic fluid power — Sample calculations for ISO 11171

Transmissions hydrauliques — Calculs des échantillons pour l'ISO 11171

Reference number
ISO/TR 6057:2023(E)

© ISO 2023



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

© ISO 2023 – All rights reserved

Contents

Page

Foreword ..... **iv**
Introduction ..... **v**
1 Scope **1**
2 Normative references **1**
3 Terms and definitions **1**
4 Example 1: Selection of threshold voltage settings **1**
5 Example 2: Evaluating data quality **3**
6 Example 3: Dilution of samples **5**
7 Example 4: Relating particle size to threshold voltage setting for particles 30 µm(c) and smaller **7**
8 Example 5: Relating particle size to threshold voltage setting for primary calibration of sizes larger than 30 µm(c) **9**
9 Example 6: Construction of calibration curve **11**
10 Example 7: Calculation of coefficient variation for volume measurement using ISO 11171:2022, Clause A.8 ..... **12**
11 Example 8: Determination of coincidence error limit **13**
12 Example 9: Determination of flow rate limits using ISO 11171:2022, Annex C **14**
13 Example 10: Determination of resolution using ISO 11171:2022, Annex D **17**
14 Example 11: Verification of counting accuracy and secondary calibration suspensions using ISO 11171:2022, Clause E.1 ..... **18**

© ISO 2023 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 131, *Fluid power systems*, Subcommittee SC 6, *Contamination control*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2023 – All rights reserved

Introduction

ISO 11171:2022, like its predecessors, retains traceability to the internationally accepted definition of a metre and reports particle size in units of μm(c). The methods for determining data acceptance criteria, coincidence error limit, working flow rate and resolution remain unchanged, but mathematical calculations and tools were first introduced in ISO 11171:2020 to ensure consistency in terms of how automatic particle counter (APC) calibration curves are created and used. For example, mathematical techniques have been introduced to determine the APC threshold settings used to obtain calibration data and a tool provided to generate calibration curves. Other mathematical equations to estimate the standard error of the calibration, to calculate normalized concentrations for diluted samples, and to calibrate at particle sizes larger than 30 μm(c) were first introduced in 2020. This document uses example calculations that fully conform to ISO 11171.

© ISO 2023 – All rights reserved

Hydraulic fluid power — Sample calculations for ISO 11171

1 Scope

This document shows how to use the normative mathematical formulae and tools of ISO 11171. Examples are used to demonstrate their use for calibrating automatic particle counters (APCs).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11171, *Hydraulic fluid power — Calibration of automatic particle counters for liquids*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11171 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

4 Example 1: Selection of threshold voltage settings

The method of selecting threshold voltages for particle sizing calibration is specified in ISO 11171:2022, 6.3, which requires that:

— a minimum of 12 different threshold settings be used to construct a calibration curve;

— the first (lowest) threshold setting, J, be 1,5 times the threshold noise level of the APC;

— the highest threshold setting, H, corresponds to a particle size of approximately 30 µm(c) or smaller for primary calibrations and corresponds to a size that does not exceed the largest reported particle size that is in conformance with ISO 11171:2022, Annex F, for secondary calibrations;

— intermediate threshold settings be logarithmically spaced such that the value of each channel is K times greater than its preceding channel, where K is a constant defined by Formula (1):

$$K = 10^{(\log H - \log J)/(G-1)} \tag{1}$$

where G is the number of threshold settings used to construct the calibration curve and is greater than or equal to 12.

This example considers an APC with eight threshold settings that can be adjusted in 1 mV increments. The threshold noise level of the APC was determined to be 5 mV and its manufacturer indicated that 30 µm(c) is expected to correspond to a threshold voltage setting of about 2 600 mV.

© ISO 2023 – All rights reserved