IEC 61010-2-032

Edition 5.0 2023-09
EXTENDED VERSION

INTERNATIONAL STANDARD

colour inside

This extended version of IEC 61010-2-032:2023 includes the content of the references made to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016

Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement



IEC 61010-2-032:2023-09 EXV(en)



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2023 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat
3, rue de Varembé
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC
The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications
The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform
The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, …). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished
Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc
If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch
Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org
The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC 61010-2-032

Edition 5.0 2023-09
EXTENDED VERSION

INTERNATIONAL STANDARD

colour inside

This extended version of IEC 61010-2-032:2023 includes the content of the references made to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016

Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 19.080

ISBN 978-2-8322-7569-6

Warning! Make sure that you obtained this publication from an authorized distributor.

® Registered trademark of the International Electrotechnical Commission

TC 66/Publication 61010-1 (2010), Third edition/I-SH 01

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 1: General requirements

INTERPRETATION SHEET 1

This interpretation sheet has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

The text of this interpretation sheet is based on the following documents:

ISH	Report on voting
66/497A/ISH	66/505/RVD

Full information on the voting for the approval of this interpretation sheet can be found in the report on voting indicated in the above table.

———————

IEC 61010-1:2010 contains a requirement in 6.8.3.1 pertaining to voltage testers for type tests as follows:

“The generator shall be able to supply a power of at least 500 VA.”

This has given rise to the following questions:

How does one interpret the requirement for voltage testers in 6.8.3.1 of IEC 61010-1:2010? Specifically, this subclause requires that “The generator shall be able to supply a power of at least 500 VA.” Does this requirement apply throughout the rated output range of the voltage tester? What is meant by the word “generator”? Is the “generator” the power supply within the voltage tester, or the voltage tester output, or something else?

Interpretation:

“A voltage tester used for type tests must be able to deliver at least 500 VA at its full-rated output voltage. It does not necessarily need to deliver 500 VA if set for lower voltages.

For example, a voltage tester that can deliver 100 mA at any test output voltage up to 5 000 V (and a current corresponding to 500 VA above 5 000 V) would meet the requirement.

The requirements for voltage testers used for routine (production line) tests are included in Annex F. The requirements of 6.8.3.1 do not apply to these voltage testers.”

February 2013 **ICS 19.080, 71.040.10**

CONTENTS

FOREWORD 11
INTRODUCTION to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016 15
INTRODUCTION 16
1 Scope and object 17
1.1 Scope 17
1.1.1 Equipment included in scope 17
1.1.2 Equipment excluded from scope 17
1.1.3 Computing equipment 17
1.2 Object 18
1.2.1 Aspects included in scope 18
1.2.2 Aspects excluded from scope 18
1.3 Verification 18
1.4 Environmental conditions 18
1.4.1 Normal environmental conditions 18
1.4.2 Extended environmental conditions 19
2 Normative references 19
3 Terms and definitions 21
3.1 Equipment and states of equipment 21
3.2 Parts and accessories 22
3.3 Quantities 23
3.4 Tests 24
3.5 Safety terms 24
3.6 Insulation 26
4 Tests 27
4.1 General 27
4.2 Sequence of tests 28
4.3 Reference test conditions 28
4.3.1 Environmental conditions 28
4.3.2 State of equipment 28
4.4 Testing in SINGLE FAULT CONDITION 30
4.4.1 General 30
4.4.2 Application of fault conditions 30
4.4.3 Duration of tests 32
4.4.4 Conformity after application of fault conditions 33
5 Marking and documentation 33
5.1 Marking 33
5.1.1 General 33
5.1.2 Identification 34
5.1.3 MAINS supply 34
5.1.4 Fuses 37
5.1.5 TERMINALS, connections and operating devices 37
5.1.6 Switches and circuit-breakers 39
5.1.7 Equipment protected by DOUBLE INSULATION or REINFORCED INSULATION 40
5.1.8 Field-wiring TERMINAL boxes 40
5.2 Warning markings 40

5.3 Durability of markings 41
5.4 Documentation 41
5.4.1 General 41
5.4.2 Equipment RATINGS 42
5.4.3 Equipment installation 42
5.4.4 Equipment operation 43
5.4.5 Equipment maintenance and service 44
5.4.6 Integration into systems or effects resulting from special conditions 44
6 Protection against electric shock 44
6.1 General 44
6.1.1 Requirements 44
6.1.2 Exceptions 44
6.2 Determination of ACCESSIBLE parts 45
6.2.1 General 45
6.2.2 Examination 45
6.2.3 Openings above parts that are HAZARDOUS LIVE 46
6.2.4 Openings for pre-set controls 46
6.3 Limit values for ACCESSIBLE parts 46
6.3.1 Levels in NORMAL CONDITION 46
6.3.2 Levels in SINGLE FAULT CONDITION 47
6.4 Primary means of protection 50
6.4.1 General 50
6.4.2 ENCLOSURES and PROTECTIVE BARRIERS 50
6.4.3 BASIC INSULATION 50
6.4.4 Impedance 50
6.5 Additional means of protection in case of SINGLE FAULT CONDITIONS 50
6.5.1 General 50
6.5.2 Not used 51
6.5.3 SUPPLEMENTARY INSULATION and REINFORCED INSULATION 51
6.5.4 PROTECTIVE IMPEDANCE 51
6.5.5 Not used 52
6.5.6 Current- or voltage-limiting device 52
6.6 Connections to external circuits 52
6.6.1 General 52
6.6.2 TERMINALS for external circuits 53
6.6.3 Circuits with TERMINALS which are HAZARDOUS LIVE 53
6.6.4 TERMINALS for stranded conductors 53
6.6.101 Measuring circuit TERMINALS 53
6.6.102 Input/output circuit leads 56
6.7 Insulation requirements 57
6.7.1 The nature of insulation 57
6.7.2 Insulation for MAINS CIRCUITS of OVERVOLTAGE CATEGORY II with a nominal supply voltage up to 300 V 59
6.7.3 Insulation for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V 63
6.8 Procedure for voltage tests 69
6.8.1 General 69
6.8.2 Humidity preconditioning 70
6.8.3 Test procedures 71

6.9 Constructional requirements for protection against electric shock 72
6.9.1 General 72
6.9.2 Insulating materials 72
6.9.3 Colour coding 72
6.10 Connection to the MAINS supply source and connections between parts of equipment 72
6.10.1 MAINS supply cords 72
6.10.2 Fitting of non-detachable MAINS supply cords 73
6.10.3 Plugs and connectors 75
6.11 Disconnection from supply source 75
6.11.1 General 75
6.11.2 Exceptions 75
6.11.3 Requirements according to type of equipment 76
6.11.4 Disconnecting devices 76
6.101 Protection against hazardous live conductors 77
6.101.1 General 77
6.101.2 Protection by a PROTECTIVE FINGERGUARD 77
6.101.3 Insulation of HAND-HELD or hand-manipulated parts for Type A current sensors 78
6.101.4 Insulation of input/output circuits 79
7 Protection against mechanical HAZARDS 79
7.1 General 79
7.2 Sharp edges 80
7.3 Moving parts 80
7.3.1 General 80
7.3.2 Exceptions 80
7.3.3 RISK assessment for mechanical HAZARDS to body parts 80
7.3.4 Limitation of force and pressure 81
7.3.5 Gap limitations between moving parts 82
7.4 Stability 84
7.5 Provisions for lifting and carrying 85
7.5.1 General 85
7.5.2 Handles and grips 85
7.5.3 Lifting devices and supporting parts 85
7.6 Wall mounting 85
7.7 Expelled parts 86
8 Resistance to mechanical stresses 86
8.1 General 86
8.2 ENCLOSURE rigidity tests 87
8.2.1 Static test 87
8.2.2 Impact test 87
8.3 Drop test 88
8.3.1 Equipment other than HAND-HELD EQUIPMENT and DIRECT PLUG-IN EQUIPMENT 88
8.3.2 HAND-HELD EQUIPMENT and DIRECT PLUG-IN EQUIPMENT 89
8.101 JAW ENDS abrasion test 89
8.102 JAW impact test 90
8.103 Pressure test at high temperature for insulation of flexible current sensors 91
8.104 Pull test for endcaps of flexible current sensors 93

8.105 Attachment of the input/output leads 93
8.105.1 General 93
8.105.2 Pull test 94
8.105.3 Flexing test 94
9 Protection against the spread of fire and arc flash 96
9.1 General 96
9.2 Eliminating or reducing the sources of ignition within the equipment 98
9.3 Containment of fire within the equipment, should it occur 98
9.3.1 General 98
9.3.2 Constructional requirements 98
9.4 Limited-energy circuit 101
9.5 Requirements for equipment containing or using flammable liquids 102
9.6 Overcurrent protection 102
9.6.1 General 102
9.6.2 PERMANENTLY CONNECTED EQUIPMENT 103
9.6.3 Other equipment 103
9.101 Protection of measuring circuits 103
9.101.1 General 103
9.101.2 Protection against mismatches of inputs and ranges 103
9.101.3 Protection against mains overvoltages 106
9.102 Prevention of hazard from arc flash and short-circuits 107
9.102.1 General 107
9.102.2 Protection against short-circuits during clamping 107
9.102.3 Protection against short-circuits in closed position 109
10 Equipment temperature limits and resistance to heat 109
10.1 Surface temperature limits for protection against burns 109
10.2 Temperatures of windings 110
10.3 Other temperature measurements 110
10.4 Conduct of temperature tests 111
10.4.1 General 111
10.4.2 Temperature measurement of heating equipment 111
10.4.3 Equipment intended for installation in a cabinet or a wall 111
10.5 Resistance to heat 112
10.5.1 Integrity of CLEARANCES and CREEPAGE DISTANCES 112
10.5.2 Non-metallic ENCLOSURES 112
10.5.3 Insulating material 112
10.5.101 Resistance to heat of current sensors 113
10.101 Other temperatures of current sensors 113
11 Protection against HAZARDS from fluids and solid foreign objects 114
11.1 General 114
11.2 Cleaning 114
11.3 Spillage 114
11.4 Overflow 114
11.5 Battery electrolyte 115
11.6 Equipment RATED with a degree of ingress protection (IP code) 115
11.6.1 General 115
11.6.2 Conditions for testing 115
11.6.3 Protection against solid foreign objects (including dust) 116
11.6.4 Protection against water 116

11.7 Fluid pressure and leakage 116
11.7.1 Maximum pressure 116
11.7.2 Leakage and rupture at high pressure 117
11.7.3 Leakage from low-pressure parts 118
11.7.4 Overpressure safety device 118
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure 119
12.1 General 119
12.2 Equipment producing ionizing radiation 119
12.2.1 Ionizing radiation 119
12.2.2 Accelerated electrons 120
12.3 Optical radiation 120
12.4 Microwave radiation 121
12.5 Sonic and ultrasonic pressure 121
12.5.1 Sound level 121
12.5.2 Ultrasonic pressure 122
12.6 Laser sources 122
13 Protection against liberated gases and substances, explosion and implosion 123
13.1 Poisonous and injurious gases and substances 123
13.2 Explosion and implosion 123
13.2.1 Components 123
13.2.2 Batteries and battery charging 123
13.2.3 Implosion of cathode ray tubes 124
14 Components and subassemblies 124
14.1 General 124
14.2 Motors 125
14.2.1 Motor temperatures 125
14.2.2 Series excitation motors 125
14.3 Overtemperature protection devices 126
14.4 Fuse holders 126
14.5 MAINS voltage selection devices 126
14.6 MAINS transformers tested outside equipment 126
14.7 Printed wiring boards 127
14.8 Circuits used to limit TRANSIENT OVERVOLTAGES 127
14.101 Probe assemblies and accessories 127
15 Protection by interlocks 128
15.1 General 128
15.2 Prevention of reactivating 128
15.3 Reliability 128
16 HAZARDS resulting from application 128
16.1 REASONABLY FORESEEABLE MISUSE 128
16.2 Ergonomic aspects 128
17 RISK assessment 129
101 Measuring circuits 129
101.1 General 129
101.2 Current sensor with an internal current transformer 130
101.3 Indicating devices 130
101.3.1 General 130

101.3.2 Battery level 130
101.3.3 Over-range indication 131
101.3.4 Permanent overvoltages 131
Annex A (normative) Measuring circuits for touch current (see 6.3) 133
Annex B (normative) Standard test fingers (see 6.2) 136
Annex C (normative) Measurement of CLEARANCES and CREEPAGE DISTANCES 138
Annex D (normative) Parts between which insulation requirements are specified (see 6.4, 6.5.3 and 6.101) 142
Annex E (informative) Guideline for reduction of POLLUTION DEGREES 146
Annex F (normative) ROUTINE TESTS 147
Annex G (informative) Leakage and rupture from fluids under pressure 150
Annex H (normative) Qualification of conformal coatings for protection against POLLUTION 155
Annex I (informative) Line-to-neutral voltages for common MAINS supply systems 158
Annex J (informative) RISK assessment 160
Annex K (normative) Insulation requirements not covered by 6.7 163
Annex L (informative) Index of defined terms 192
Annex AA (normative) MEASUREMENT CATEGORIES 194
Annex BB (informative) HAZARDS pertaining to measurements performed in certain environments 197
Annex CC (informative) 4 mm "banana" TERMINALS 200
Annex DD (informative) Flowchart for insulation according to the type of circuit 202
Annex EE (informative) Determination of CLEARANCES for Table 101 205
Annex FF (normative) Type of current sensors 206
Annex GG (normative) CLAMP MULTIMETER 208
Bibliography 210

Figure 1 – Measurements through openings in ENCLOSURES 46
Figure 2 – Maximum duration of short-term ACCESSIBLE voltages in SINGLE FAULT CONDITION (see 6.3.2 a)) 48
Figure 3 – Capacitance level versus voltage in NORMAL CONDITION and SINGLE FAULT CONDITION (see 6.3.1 c) and 6.3.2 c)) 49
Figure 4 – Acceptable arrangement of protective means against electric shock 51
Figure 6 – Distance between conductors on an interface between two layers 61
Figure 7 – Distance between adjacent conductors along an interface of two inner layers 62
Figure 8 – Distance between adjacent conductors located between the same two layers 63
Figure 9 – Detachable MAINS supply cords and connections 73
Figure 101 – CLEARANCE between the PROTECTIVE FINGERGUARD to the JAWS and to the HAZARDOUS LIVE conductor 78
Figure 10 – Impact test using a sphere 88
Figure 102 – Abrasion test of the JAW ENDS 90
Figure 103 – Impact points for JAW impact test 91
Figure 104 – Indentation device 92
Figure 105 – Example of a current sensor with strain reliefs 94

Figure 106 – Flexing test 95
Figure 11 – Flow chart to explain the requirements for protection against the spread of fire 97
Figure 12 – Baffle 100
Figure 13 – Area of the bottom of an ENCLOSURE to be constructed as specified in 9.3.2 c) 1) 100
Figure 107 – Test probe to check protection against short-circuits 108
Figure 108 – Use of the test probe of Figure 107 108
Figure 14 – Ball-pressure test apparatus 113
Figure 16 – Ratio between test pressure and maximum working pressure 118
Figure 15 – Flow chart for conformity options 14.1 a), b), c) and d) 125
Figure A.1 – Measuring circuit for a.c. with frequencies up to 1 MHz and for d.c. 133
Figure A.2 – Measuring circuits for sinusoidal a.c. with frequencies up to 100 Hz and for d.c. 134
Figure A.3 – Current measuring circuit for electrical burns 135
Figure A.4 – Current measuring circuit for wet contact 135
Figure B.1 – Rigid test finger 136
Figure B.2 – Jointed test finger 137
Figure C.1 – Examples of methods of measuring CLEARANCES and CREEPAGE DISTANCES 141
Figures D.1a) to d) – Protection between HAZARDOUS LIVE circuits and ACCESSIBLE parts 143
Figures D.1e) to h) – Protection between HAZARDOUS LIVE circuits and circuits with ACCESSIBLE external TERMINALS 143
Figures D.2 a) and D.2 b) – Protection between a HAZARDOUS LIVE internal circuit and an ACCESSIBLE part which is not bonded to other ACCESSIBLE parts 144
Figures D.2 c) and D.2 d) – Protection between a HAZARDOUS LIVE primary circuit and circuits which have ACCESSIBLE external TERMINALS 144
Figure D.3 – Protection of external ACCESSIBLE TERMINALS of two HAZARDOUS LIVE circuits 144
Figure D.101 – Parts of current sensors (see also Table D.101) 145
Figure G.1 – Conformity verification process (see G.2) 151
Figure H.1 – Test sequence and conformity 157
Figure J.1 – Iterative process of RISK assessment and RISK reduction 160
Figure J.2 – RISK reduction 161
Figure K.1 – Distance between conductors on an interface between two layers 167
Figure K.2 – Distance between adjacent conductors along an interface of an inner layer 168
Figure K.3 – Distance between adjacent conductors located between the same two layers 169
Figure K.4 – Example of recurring peak voltage 182
Figure K.101 – CLEARANCES between the conductor, the input/output circuit and the enclosure of a Type A current sensor 184
Figure K.102 – Test circuit for evaluation of TRANSIENT OVERVOLTAGE attenuation 186
Figure AA.1 – Example to identify the locations of MEASUREMENT CATEGORIES 195
Figure CC.1 – Recommended dimensions of 4 mm TERMINALS 201
Figure DD.1 – Requirements for CLEARANCE, CREEPAGE DISTANCE and solid insulation 204

Figure FF.1 – Examples of current sensors and their parts 207
Figure GG.1 – Examples of CLAMP MULTIMETERS 208

Table 1 – Symbols 36
Table 101 – CLEARANCES for unmated measuring circuit TERMINALS 54
Table 3 – Multiplication factors for CLEARANCES of equipment RATED for operation at altitudes up to 5 000 m 58
TABLE 4 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V 60
Table 5 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V 60
Table 6 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V 65
Table 7 – CREEPAGE DISTANCES for secondary circuits 66
Table 8 – Minimum values for distance or thickness (see 6.7.3.4.2 to 6.7.3.4.4) 68
Table 9 – Distances between TERMINALS and foil 69
Table 10 – Correction factors according to test site altitude for test voltages for CLEARANCES 70
Table 11 – Values for physical tests on cord anchorages 74
Table 12 – Protective measures against mechanical HAZARDS to body parts 81
Table 13 – Minimum maintained gaps to prevent crushing for different body parts 83
Table 14 – Maximum gaps to prevent access for different body parts 83
Table 15 – Impact energy levels, test height and corresponding IK codes 88
Table 102 – Energy level for JAW impact test 90
Table 103 – Pull forces for endcaps of flexible current sensors 93
Table 104 – Pull forces for input/output lead attachment test 96
Table 16 – Acceptable perforation of the bottom of an ENCLOSURE 99
Table 17 – Limits of maximum available current 101
Table 18 – Values for overcurrent protection devices 102
Table 105 – Thickness of the test probe of Figure 107 and test voltages 109
Table 19 – Surface temperature limits in NORMAL CONDITION 110
Table 20 – Maximum temperatures for insulation material of windings 110
Table 22 – Lamp or lamp systems considered photobiologically safe 121
Table 23 – Lamp or lamp systems considered photobiologically safe under certain conditions 121
Table 21 – Impulse voltages for OVERVOLTAGECATEGORY II 127
Table C.1 – Dimensions of X 138
Table D.101 – Insulation requirements for current sensors 145
Table E.1 – Environmental situations 146
Table E.2 – Reduction of POLLUTION DEGREES 146
Table F.1 – Test voltages for ROUTINE TESTS of MAINS CIRCUITS 148
Table G.1 – Test pressures for equipment with pressures above 14 Mpa 153
Table H.1 – Test parameters, test conditions and test procedures 156
Table I.1 – Line-to-neutral voltages for common MAINS supply systems 158
Table J.1 – Severity of harm 162

Table J.2 – Probability of harm 162
Table J.3 – RISK category 162
Table K.1 – Multiplication factors for CLEARANCES for equipment RATED for operation at altitudes up to 5 000 m 164
Table K.2 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY II above 300 V 164
Table K.3 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY III 165
Table K.4 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY IV 165
Table K.5 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY II above 300 V 166
Table K.6 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY III 166
Table K.7 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY IV 166
Table K.8 – Test voltages for testing long-term stress of solid insulation in MAINS CIRCUITS 167
Table K.9 – Minimum values for distance or thickness of solid insulation 168
Table K.10 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY II above 300 V 171
Table K.11 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY III 172
Table K.12 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY IV 173
Table K.13 – CREEPAGE DISTANCES for secondary circuits 174
Table K.14 – Minimum values for distance or thickness (see K.2.4.2 to K.2.4.4) 176
Table K.15 – CLEARANCE values for the calculation of K.3.2 179
Table K.16 – Test voltages based on CLEARANCES 180
Table K.17 – CLEARANCES for BASIC INSULATION in circuits having recurring peak voltages or WORKING VOLTAGES with frequencies above 30 kHz 182
Table K.101 – Impulse voltages for circuits connected to MAINS 186
Table K.102 – CLEARANCES for measuring circuits RATED for MEASUREMENT CATEGORIES 187
Table K.103 – Impulse test voltages for testing electric strength of solid insulation for measuring circuits RATED for MEASUREMENT CATEGORIES 188
Table K.104 – a.c. test voltages for testing electric strength of solid insulation for measuring circuits RATED for MEASUREMENT CATEGORIES 189
Table K.105 – Minimum values for distance or thickness of solid insulation for measuring circuits RATED for MEASUREMENT CATEGORIES 190
Table AA.1 – Characteristics of MEASUREMENT CATEGORIES 196
Table EE.1 – CLEARANCE values for Table 101 205

INTERNATIONAL ELECTROTECHNICAL COMMISSION

_____

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 1: General requirements

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This extended version (EXV) of the official IEC Standard provides the user with the comprehensive content of the Standard.

IEC 61010-2-032:2023 EXV includes the content of IEC 61010-2-032:2023, and the references made to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016.

The specific content of IEC 61010-2-032:2023 is displayed on a blue background.

IEC 61010-2-032 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment. It is an International Standard.

This fifth edition cancels and replaces the fourth edition published in 2019. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

a) In 1.1.1, definitions of current sensor types have been moved to a new Annex FF;
b) Clause 2, all normative references have been dated and new normative references have been added;
c) 3.2.103, a new definition PROTECTIVE FINGERGUARD has been added which replaces the previous definition of PROTECTIVE BARRIER;
d) 4.4.2.101 is a new subclause about surge protective devices;
e) in 5.1.5.101.2, minimum RATINGS for voltage of measuring TERMINALS are required;
f) Subclause 6.5.1 has been modified;
g) Subclause 6.5.5 is no longer used;
h) Subclause 6.6.101 modifies 6.6.101 and 6.6.102 of previous edition:
    1) in 6.6.101.1, insulating material of group I may be allowed for determination of CREEPAGE DISTANCES of measuring circuit TERMINALS;
    2) in 6.6.101.2, CLEARANCES and CREEPAGE DISTANCES up to 3 000 V for measuring circuit TERMINALS in unmated position have been defined;
    3) in 6.6.101.3, requirements for measuring circuit TERMINALS in partially mated position have been specified;
    4) in 6.6.101.4, requirements for measuring circuit TERMINALS in mated position have been specified;
    5) Subclause 6.6.101.5 replaces 6.6.102;
i) Subclause 6.6.102 replaces 6.101 of previous edition with modifications;
j) Subclause 6.101.2 replaces 6.9.101.1 of previous edition with modifications;
k) Subclause 6.101.3 replaces 6.9.101.2 of previous edition with modifications;
l) Subclause 6.101.4 replaces 6.9.102 of previous edition with modifications;
m) in 8.101, JAW ENDS abrasion test has been modified;
n) 8.105 is a new subclause for input/output leads attachment has been added;
o) in 9.101.2, relocation of 101.3 of previous edition;
p) in 9.101.3, relocation of 101.4 of previous edition, extension to MEASUREMENT CATEGORY II and reference to IEC 61000-4-5 for tests;
q) Table 104 has been replaced by Table K.101;
r) in 9.102, relocation of Clause 102 of previous edition;
s) in 14.101, relocation of 14.102. Subclause 14.101 of previous edition has been deleted;
t) 101.3 is a new subclause for protections against HAZARD occurring from reading a voltage value in replacement of Clause EE.5 of previous edition;
u) in Table D.101, transients are disregarded for insulation between JAW ENDS and input/output circuits;
v) in Clause F.101, test voltages for routine test of JAWS have been modified;
w) in K.2.1, another method for determination of CLEARANCES of secondary circuits is proposed;
x) in K.3.2, new Table K.15 and Table K.16 for CLEARANCE calculation;

y) K.3.101 is a new clause;

z) Clause K.4, redraft of the clause to propose a method for determination of U_t for circuits which reduce TRANSIENT OVERVOLTAGES;

aa) Table K.101 replaces Table 104;

bb) Subclause K.101.4 has been reviewed and tables and tests for solid insulation have been modified;

cc) Table K.104 of previous edition has been deleted;

dd) Annex AA: Figure AA.1 has been redesigned;

ee) Annex EE: addition of a new informative annex for determination of CLEARANCES for Table 101;

ff) Annex GG: this annex was Annex EE of previous edition and the current sensor type of a CLAMP MULTIMETER is type A or type B.

The text of this International Standard is based on the following documents:

Draft	Report on voting
66/788A/FDIS	66/798/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

A list of all parts of the IEC 61010 series, under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

This document is to be used in conjunction with IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016.

This document supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement*.

Where a particular subclause of IEC 61010-1 is not mentioned in this document, that subclause applies as far as is reasonable. Where this document states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in IEC 61010-1 should be adapted accordingly.

In this standard:

a) the following print types are used:
   - requirements: in roman type;
   - NOTES: in small roman type;
   - *conformity and tests: in italic type;*
   - terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS;

b) subclauses, figures, tables and notes which are additional to those in IEC 61010-1 are numbered starting from 101. Additional annexes are lettered starting from AA and additional list items are lettered from aa).

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016

This Interational Standard specifies the safety requirements that are generally applicable to all equipment within its scope. For certain types of equipment, these requirements will be supplemented or modified by the special requirements of one, or more than one, particular part 2 of the standard which must be read in conjunction with the part 1 requirements.

INTRODUCTION

IEC 61010-1 specifies the safety requirements that are generally applicable to all equipment within its scope. For certain types of equipment, the requirements of IEC 61010-1 and its amendment will be supplemented or modified by the special requirements of one or more standard from the IEC 61010-2 series which is/are read in conjunction with the requirements of IEC 61010-1.

1) IEC 61010-2-030:2023 specifies the safety requirements for equipment with testing or measuring circuits which are connected for test or measurement purposes to devices or circuits outside the measurement equipment itself.
2) This document specifies the safety requirements for hand-held and hand-manipulated current sensors for measuring, detecting or injecting current, or indicating current waveforms on circuits without physically opening the current path of the current path of the circuit being measured.

Most of the requirements of IEC 61010-2-030:2023 have been included in this document. Equipment within the scopes of both IEC 61010-2-030:2023 and this document is considered to be covered by the requirements of this document.

However, for current sensors in combined equipment with protective bonding and automatic disconnection of the supply, IEC 61010-2-030:2023 and this document are read in conjunction.

3) IEC 61010-2-033:2023 specifies the safety requirements for hand-held multimeters and other meters for domestic and professional use, capable of measuring mains voltage, intended to measure voltage and other electrical quantities such as resistance or current.

All relevant requirements of IEC 61010-2-030 have been included in IEC 61010-2-033:2023.

4) IEC 61010-2-034:2023 specifies the safety requirements for measurement equipment for insulation resistance and test equipment for electric strength which are connected to units, lines or circuits for test or measurement purposes.

All relevant requirements of IEC 61010-2-030:2023 have been included in IEC 61010-2-034:2023. However, for equipment within the scope of this document and IEC 61010-2-034:2023, these standards are read in conjunction.

IEC 61010-031 specifies the safety requirements for hand-held and hand-manipulated probe assemblies and their related accessories intended to be used in particular with equipment in the scope of IEC 61010-2-030, this document, IEC 61010-2-033 and IEC 61010-2-034. These probe assemblies are for non-contact or direct electrical connection between a part and electrical test and measurement equipment. They may be fixed to the equipment or be detachable accessories for the equipment.

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 1: General requirements

1 Scope and object

1.1 Scope

1.1.1 Equipment included in scope

This document specifies safety requirements for HAND-HELD and hand-manipulated current sensors intended for measuring, detecting or injecting current, or indicating current waveforms on circuits without physically opening the current path of the circuit being measured.

These current sensors are hand-manipulated before and/or after a test or measurement, but are not necessarily HAND-HELD during the test or measurement. They can be stand-alone current sensors or accessories to other equipment or parts of combined equipment. These include measurement circuits which are part of electrical test and measurement equipment, laboratory equipment, or process control equipment.

NOTE 1 Combined equipment is equipment that is electrically connected to a current sensor by means of a permanent connection which can be detached only by the use of a TOOL.

NOTE 2 Some current sensors are also known as current clamps, CLAMP MULTIMETERS and current probes.

The types of current sensors covered by this document are defined in Annex FF.

1.1.2 Equipment excluded from scope

This standard does not apply to equipment within the scope of:

a) IEC 60065 (Audio, video and similar electronic apparatus);
b) IEC 60204 (Safety of machinery – Electrical equipment of machines);
c) IEC 60335 (Household and similar electrical appliances);
d) IEC 60364 (Electrical installations of buildings);
e) IEC 60439 (Low-voltage switchgear and controlgear assemblies);
f) IEC 60601 (Medical electrical equipment);
g) IEC 60950 (Information technology equipment including electrical business equipment, except as specified in 1.1.3);
h) IEC 61558 (Power transformers, power supply units and similar);
i) IEC 61010-031 (Hand-held probe assemblies);
j) IEC 61243-3 (Live working – Voltage detectors – Part 3: Two-pole low-voltage type).

This document does not apply to current sensors used as FIXED EQUIPMENT.

1.1.3 Computing equipment

This standard applies only to computers, processors, etc. which form part of equipment within the scope of this standard or are designed for use exclusively with the equipment.

NOTE Computing devices and similar equipment within the scope of IEC 60950 and conforming to its requirements are considered to be suitable for use with equipment within the scope of this standard. However,