IEC 62618

Edition 2.0 2022-11

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Radiation protection instrumentation – Spectroscopy-based alarming personal radiation detectors (SPRD) for the detection of illicit trafficking of radioactive material

Instrumentation pour la radioprotection – Détecteurs individuels spectroscopiques d'alarme aux rayonnements (SPRD) pour la détection du trafic illicite de matières radioactives



IEC 62618:2022-11(en-fr)



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2022 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembé
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC
The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications
The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform
The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, …). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished
Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc
If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch
Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org
The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC
La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC
Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC - webstore.iec.ch/advsearchform
La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, …). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished
Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc
Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch
Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org
Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 300 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 19 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

IEC 62618

Edition 2.0 2022-11

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Radiation protection instrumentation – Spectroscopy-based alarming personal radiation detectors (SPRD) for the detection of illicit trafficking of radioactive material

Instrumentation pour la radioprotection – Détecteurs individuels spectroscopiques d'alarme aux rayonnements (SPRD) pour la détection du trafic illicite de matières radioactives

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 13.280

ISBN 978-2-8322-6050-0

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

® Registered trademark of the International Electrotechnical Commission
Marque déposée de la Commission Electrotechnique Internationale

CONTENTS

FOREWORD 4
INTRODUCTION 6
1 Scope 7
2 Normative references 7
3 Terms and definitions, abbreviated terms and symbols, quantities and units 8
3.1 Terms and definitions 8
3.2 Abbreviated terms and symbols 10
3.3 Quantities and units 10
4 General test procedure 11
4.1 Standard test conditions 11
4.2 Uncertainties 11
4.3 Statistical fluctuations 11
4.4 Background radiation during testing 12
4.5 Operating parameters and set up 12
4.6 Radiation sources 12
4.7 Special nuclear material (SNM) and depleted uranium (DU) sources 12
4.8 Speed of moving sources including scaling 12
4.9 Functionality test and test acceptance range requirements 13
4.10 Neutron measurement requirements 15
5 General requirements 16
5.1 Basic information 16
5.2 Mechanical characteristics 16
5.3 Data output 17
5.4 User indications 17
5.5 Markings 17
5.6 Alarms 18
6 Radiation detection requirements 18
6.1 False alarm rate 18
6.2 Gamma alarm 18
6.3 Personal protection alarm 19
6.4 Relative intrinsic error 19
6.5 Detection of gradually increasing radiation levels 20
6.6 Over-range 20
6.7 Detection of neutrons (if provided) 20
6.8 Gamma response of neutron detector (if provided) 21
6.9 Identification of single radionuclides 22
6.10 Simultaneous radionuclide identification 23
6.11 Low-exposure rate identification 23
6.12 Over range characteristics for identification 23
7 Environmental requirements 24
7.1 Ambient temperature 24
7.2 Temperature shock 24
7.3 Relative humidity 24
7.4 Dust and moisture protection 25
8 Mechanical requirements 25
8.1 Vibration 25

8.2 Microphonics/Impact 26
8.3 Drop 26
9 Electromagnetic requirements 26
9.1 Electrostatic discharge (ESD) 26
9.2 Radio frequency (RF) 27
9.3 Radiated RF emissions 27
9.4 Magnetic fields 27
10 Documentation 28
10.1 Operation and maintenance manual 28
10.2 Test certificate 28
10.3 Declaration of conformity 28
Bibliography 29

Table 1 – Overview of IEC radiation protection instrumentation standards 6
Table 2 – Standard test conditions 11
Table 3 – Test results analysis 15
Table 4 – Identification acceptance criteria[1] 22

INTERNATIONAL ELECTROTECHNICAL COMMISSION

_____

RADIATION PROTECTION INSTRUMENTATION – SPECTROSCOPY-BASED ALARMING PERSONAL RADIATION DETECTORS (SPRD) FOR THE DETECTION OF ILLICIT TRAFFICKING OF RADIOACTIVE MATERIAL

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62618 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation. It is an International Standard.

This second edition cancels and replaces the first edition published in 2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

a) making the standard consistent with the new standards for detection of illicit trafficking of radioactive material (see the Introduction);
b) creating unformed functionality test for all environmental, electromagnetic and mechanical tests and a requirement for the coefficient of variation of each nominal mean reading;
c) reference to IEC 62706 for the environmental, electromagnetic and mechanical test conditions;
d) adding information regarding climatic exposures.

The text of this International Standard is based on the following documents:

Draft	Report on voting
45B/1011/FDIS	45B/1017/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

It is important to detect illicit and inadvertent movement of radioactive materials in the form of radiation sources and contaminated metallurgical scrap. Radioactive sources out of regulatory control, so-called “orphan sources”, have frequently caused serious radiation exposures and widespread contamination. Although illicit trafficking of nuclear and other radioactive materials is not a new problem, concern about a nuclear “black market” has increased particularly in view of its terrorist potential.

In response to the technical policy of the International Atomic Energy Agency (IAEA), the World Customs Organization (WCO), and the International Criminal Police Organization (Interpol) related to the detection and identification of special nuclear materials and security trends, nuclear instrumentation companies have developed and manufactured radiation instrumentation to assist in the detection of illicit movement of radioactive and special nuclear materials. This type of instrumentation is widely used for security purposes at nuclear facilities, border control checkpoints, and international seaports and airports.

To ensure that measurement results made at different locations are consistent, it is imperative that radiation instrumentation be designed to specifications based upon agreed performance requirements. IEC standards have been developed to establish performance requirements for personal radiation detectors, radiation portal monitors, highly sensitive gamma and neutron detection systems, spectrometric personal radiation detectors, and backpack-based radiation detection and identification systems. Table 1 contains a list of those standards.

Table 1 – Overview of IEC radiation protection instrumentation standards

Type of instrumentation	IEC number	Title of the standard
Body-worn	**62401**	Radiation protection instrumentation – Alarming Personal Radiation Devices (PRDs) for the detection of illicit trafficking of radioactive material
	62618	Radiation protection instrumentation – Spectroscopy-Based Alarming Personal Radiation Detectors (SPRD) for the detection of illicit trafficking of radioactive material
	62694	Radiation protection instrumentation – Backpack-type radiation detector (BRD) for the detection of illicit trafficking of radioactive material
Portable or hand-held	**62327**	Radiation protection instrumentation – Hand-held instruments for the detection and identification of radionuclides and for the estimation of ambient dose equivalent rate from photon radiation
	62533	Radiation protection instrumentation – Highly sensitive hand-held instruments for photon detection of radioactive material
	62534	Radiation protection instrumentation – Highly sensitive hand-held instruments for neutron detection of radioactive material
Portal	**62244**	Radiation protection instrumentation – Installed radiation portal monitors (RPMs) for the detection of illicit trafficking of radioactive and nuclear materials
	62484	Radiation protection instrumentation – Spectrometric radiation portal monitors (SRPMs) used for the detection and identification of illicit trafficking of radioactive material
Data format	**62755**	Radiation protection instrumentation – Data format for radiation instruments used in the detection of illicit trafficking of radioactive materials
Mobile system	**63121**	Radiation protection instrumentation – Vehicle-mounted mobile systems for the detection of illicit trafficking of radioactive materials

RADIATION PROTECTION INSTRUMENTATION – SPECTROSCOPY-BASED ALARMING PERSONAL RADIATION DETECTORS (SPRD) FOR THE DETECTION OF ILLICIT TRAFFICKING OF RADIOACTIVE MATERIAL

1 Scope

This document applies to Spectroscopy-based alarming Personal Radiation Detectors (SPRD). SPRDs detect and identify gamma radiation and may detect neutron radiation. SPRDs can be worn on a belt or in a pocket to alert the wearer of the presence of a radiation source. SPRDs provide search, similar to that of a Personal Radiation Device (PRD), and identification capability to identify radiation sources. They can discriminate between alarms caused by Naturally Occurring Radioactive Materials (NORM) or medical radionuclides and alarms from industrial sources or Special Nuclear Material (SNM).

This document establishes performance requirements and specifies general characteristics, general test conditions, radiological, climatic, mechanical, and electromagnetic characteristics. This document also provides test methods that are used to determine if an SPRD meets the stated requirements.

This document does not apply to the performance of radiation protection instrumentation which is covered in IEC 61526 and IEC 60846-1. SPRDs are not intended for accurate measurement of personal ($H_p(10)$) or ambient ($H^*(10)$) dose equivalent (rate).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-395:2014, *International Electrotechnical Vocabulary (IEV) – Part 395: Nuclear instrumentation – Physical phenomena, basic concepts, instruments, systems, equipment and* detectors
IEC 60050-395:2014/AMD1:2016
IEC 60050-395:2014/AMD2:2020

IEC 60079-11, *Explosive atmospheres – Part 11: Equipment protection by intrinsic safety "i"*

IEC 62706, *Radiation protection instrumentation – Recommended climatic, electromagnetic and mechanical performance requirements and methods of tests*

IEC 62755, *Radiation protection instrumentation – Data format for radiation instruments used in the detection of illicit trafficking of radioactive materials*

IEC TR 62971:2015, *Radiation instrumentation – Radiation sources used in illicit trafficking detection standards – Guidance and recommendations*

UL 913, *Standard for Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, III, Division 1, Hazardous (Classified) Locations*

ICRU report 39, *Determination of Dose Equivalents Resulting from External Radiation Sources*

ICRU report 47, *Measurement of Dose Equivalents from External Photon and Electron Radiations*

3 Terms and definitions, abbreviated terms and symbols, quantities and units

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60050-395, as well as the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at https://www.electropedia.org/
- ISO Online browsing platform: available at https://www.iso.org/obp

3.1.1
acceptable identification
correct identification
when an instrument correctly identifies only the radionuclides present

3.1.2
accuracy of measurement
closeness of the agreement between the result of a measurement and the conventionally true value of the measurand

Note 1 to entry: "Accuracy" is a quantitative concept.

Note 2 to entry: The term precision should not be used for "accuracy".

[SOURCE: IEC 60050-395:2014/AMD2:2020, 395-16-14]

3.1.3
alarm
audible, visual, or other signal activated when the instrument reading exceeds a pre-set value or falls outside of a pre-set range

3.1.4
alarm criteria
condition that causes an instrument to alarm

3.1.5
background radiation level
radiation field in which there are no external radioactive sources present other than those in the natural background at the location of the measurements

3.1.6
confidence indication
indication provided by the instrument to assess the reliability assigned to the validity of the identification. For each identified radionuclide, the instrument indicates the likelihood of its correct identification.

3.1.7
coefficient of variation
ratio of the standard deviation σ to the arithmetic mean $\overline{x}$ of a set of n measurements x_i given by the following formula:

$$COV = \frac{s}{\overline{x}} = \frac{1}{\overline{x}}\sqrt{\frac{1}{n-1}\sum_{i=1}^{n}\left(x_i - \overline{x}\right)^2}$$

3.1.8
conventionally true value of a quantity
value attributed to a particular quantity and accepted, sometimes by convention, as having an uncertainty appropriate for a given purpose

Note 1 to entry: "Conventionally true value of a quantity" is sometimes called assigned value, best estimate of the value, conventional value or reference value.

Note 2 to entry: A conventionally true value is, in general, regarded as sufficiently close to the true value for the difference to be insignificant for the given purpose. For example, a value determined from a primary or secondary standard or by a reference instrument, may be taken as the conventionally true value.

3.1.9
false alarm
alarm not caused by an increase in radiation level over background conditions

3.1.10
functionality test
procedure to measure potential changes in the instrument response, such as drift in energy calibration or sensitivity

3.1.11
influence quantity
quantity that is not the measurand but that affects the result of the measurement

Note 1 to entry: For example, temperature of a micrometer used to measure length.

3.1.12
precision
degree to which repeated measurements under unchanged conditions show the same result (also called reproducibility or repeatability)

3.1.13
radioactive material
material containing one or more constituents exhibiting radioactivity

Note 1 to entry: For the purpose of this document, radioactive material includes special nuclear material.

3.1.14
reference point of an instrument
mark on the equipment at which the instrument is positioned for the purpose of calibration

Note 1 to entry: The point from which the distance to the source is measured.

Note 2 to entry: The reference point for calibration is also used as reference point for testing.

3.1.15
reference source
radioactive secondary standard source for use in calibration of the measuring instrument

Note 1 to entry: In this document, reference sources for calibration are used for testing.

3.1.16
relative intrinsic error
difference between the instrument's reading, M, and the conventionally true value, CTV, of the quantity being measured divided by the conventionally true value when subjected to a specified reference quantity under specified reference conditions

$$\varepsilon_{REL} = \frac{M - CTV}{CTV}$$

3.1.17
safety alarm
audible and visual signal for detection of high radiation levels, which requires immediate radiation safety measures

3.1.18
source indication alarm
audible and/or visual signal to indicate the presence of a radiation source

3.1.19
standard test conditions
range of values of a set of influence quantities under which a test, calibration or measurement of response is carried out

3.1.20
performance test
environmental, mechanical or electrical test taken from IEC 62706

[SOURCE: IEC 60050-395:2014/AMD2:2020, 395-14-9]

3.2 Abbreviated terms and symbols

CISPR	Comité International Spécial des Perturbations Radioélectriques (Special International Committee on Radio Interference)
CLYC	Cs_2LiYCl_6:Ce
COV	coefficient of variation
CZT	cadmium zinc telluride
DU	depleted uranium
ESD	electrostatic discharge
HDPE	highly-density polyethylene
HEU	highly-enriched uranium
ICRU	International commission on radiation units and measurements
NORM	naturally occurring radioactive material
PMMA	polymethyl methacrylate
RF	radio frequency
SNM	special nuclear material
SPRD	spectroscopy based personal radiation detector
WGPu	weapons-grade plutonium

3.3 Quantities and units

In this document, units of the International System (SI) are used[1]. The definitions of radiation quantities are given in IEC 60050-395.

[1] International Bureau of Weights and Measures: The International System of Units, 8th edition, 2006.

The following units may also be used:

– for energy: electron-volt (symbol: eV), 1 eV = 1,602 × 10^{-19} J;
– for time: years (symbol: y), days (symbol: d), hours (symbol: h), minutes (symbol: min);
– for temperature: degrees Celsius (symbol: °C), 0 °C = 273,15 K.

Multiples and submultiples of SI units are used, when practicable, according to the SI system.

4 General test procedure

4.1 Standard test conditions

Except where otherwise specified, the tests described in this document should be performed under the standard test conditions given in Table 2. The ambient temperature, relative humidity, and atmospheric pressure shall be recorded during testing.

Table 2 – Standard test conditions

Influence quantity	Standard test conditions
Ambient temperature	18 °C to 25 °C
Relative humidity	≤ 75 %
Atmospheric pressure	70 kPa to 103,3 kPa
Gamma radiation background	Ambient dose equivalent rate less than or equal to 0,15 μSv•h^{-1} Natural conditions without the presence of man-made emitters
Neutron background	Natural conditions without the presence of man-made emitters

NOTE SPRDs are typically used in non-radiological areas, e.g., shipping ports and border locations. Man-made radiological materials such as radiation sources are not expected to be present in these areas. Non-radiological areas are expected to be used when testing SPRDs.

4.2 Uncertainties

Unless otherwise stated for a specific quantity, the uncertainties for measurable quantities should not exceed 15 % with a coverage factor of k = 1, except for background radiation measurements that may exceed this value.

4.3 Statistical fluctuations

For tests involving the use of radioactive sources to verify susceptibility to a climatic, electromagnetic, or mechanical condition (Clauses 7, 8, and 9), the radiation field produced by the sources to verify an SPRD's response shall be adjusted to reduce the magnitude of the statistical fluctuations.

For measurements without sources (i.e., at the level of background radiation), the SPRD is observed in order to verify that alarms and spurious indications are not produced by an influence quantity (e.g., temperature, humidity, RF, impact, vibration), as readings are expected to display large fluctuations.