International Standard

ISO 15747

Plastic containers for intravenous injections

Récipients en plastique pour injections intraveineuses

Fourth edition
2026-05

Reference number
ISO 15747:2026(en)

© ISO 2026



COPYRIGHT PROTECTED DOCUMENT

© ISO 2026

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword ... **iv**
1 Scope ... **1**
2 Normative references ... **1**
3 Terms and definitions ... **1**
4 Requirement ... **3**
4.1 Physical requirements ... 3
4.1.1 General ... 3
4.1.2 Manufacturing process compatibility ... 3
4.1.3 Resistance to temperature, pressure and leakage ... 3
4.1.4 Resistance to dropping ... 3
4.1.5 Transparency ... 3
4.1.6 Water vapour permeability ... 3
4.1.7 Particulate contamination ... 3
4.1.8 Cover ... 3
4.1.9 Access port — Penetration ability of the insertion point ... 4
4.1.10 Access port — Adhesion strength of the infusion device and impermeability of the insertion point ... 4
4.1.11 Access port — Tightness of the injection point ... 4
4.1.12 Hanger ... 4
4.1.13 Identification ... 4
4.2 Chemical requirements ... 4
4.2.1 General ... 4
4.2.2 Requirements for the raw container or the sheeting ... 4
4.2.3 Requirements for the test fluid ... 5
4.3 Biological requirements ... 5
4.3.1 General ... 5
4.3.2 Impermeability for microorganisms ... 5
4.3.3 Migration ... 5
5 Application of tests ... **5**
Annex A (normative) **Physical tests** ... **6**
Annex B (normative) **Chemical tests** ... **9**
Annex C (normative) **Biological tests** ... **12**
Annex D (normative) **Reference spike** ... **14**
Annex E (informative) **Rationale, guidance and history of the development of this document** ... **16**
Annex F (informative) **Sustainability** ... **19**
Annex G (informative) **Attributive and variable testing** ... **21**
Bibliography ... **22**

© ISO 2026 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use,* in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices,* in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This fourth edition cancels and replaces the third edition (ISO 15747:2018), which has been technically revised.

The main changes are as follows.

— The term “cannula” has been replaced by the better suited term “needle” throughout the document. Per common understanding, cannulas are flexible, while needles, such as 23G needles called for in A.10, as well as other so-called “transfer needles/devices” used to add medication through an injection point, whether they are sharp or blunt, need to be rigid to pierce said injection port and are typically made of stainless steel or moulded thermoplastic.

— Addition of subclauses 4.1.1, 4.2.1 and 4.3.1 to highlight the normative nature of the introductory subclauses A.1, A.2, B.1 and C.1, as they contain important information to properly conduct the tests. Those introductory clauses were not directly referenced in any requirement since they don’t describe test methods related to said requirements. Addition of those three subclauses led to the renumbering of all other subclauses of Clause 4.

— Usage of terms “procedure” and “method” clarified: “method” now used as way to perform a test, while “procedure” denotes a process to reach a certain state (e.g. thinking process or working process).

— Addition of a new Annex E (Rationale and guidance), to provide explanations about the history of the development of the standard and to summarize the different arguments discussed within ISO/TC 76 during the elaboration of the document.

— Addition of a new Annex F (Sustainability) and a new Annex G (Attributive and variable testing).

— Addition of references to pharmacopoeias pertaining to chemical requirements, in 4.2.

© ISO 2026 – All rights reserved

— Replacement of reference to ISO 2768-2 (withdrawn) by ISO 22081 in Figure D.1, Annex D and Clause 2. Consequently, Figure D.1 has been modified (addition of reference datums, etc.)

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2026 – All rights reserved

Plastic containers for intravenous injections

1 Scope

This document specifies requirements to the safe handling and the physical, chemical and biological testing of plastic containers for parenterals.

This document is applicable to plastic containers for parenterals having one or more chambers and having a total nominal capacity in the range of 50 ml to 5 000 ml such as film bags or blow-moulded plastic bottles for direct administration of infusion (injection) solutions.

NOTE 1 In some countries, national or regional pharmacopoeias or other government regulations are legally binding, and these requirements take precedence over this document.

NOTE 2 Annex E provides explanations about the history of the development of the standard and summarises the different arguments discussed within ISO/TC 76 during the elaboration of the document.

NOTE 3 Annex F provides recommendations regarding sustainability.

NOTE 4 Annex G provides information on attributive and variable testing.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2768-1, *General tolerances — Part 1: Tolerances for linear and angular dimensions without individual tolerance indications*

ISO 8536-4, *Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Requirements and general principles for the evaluation of biological safety within a risk management process*

ISO 10993-5, *Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity*

ISO 15510, *Stainless steels — Chemical composition*

ISO 22081, *Geometrical product specifications (GPS) — Geometrical tolerancing — General geometrical specifications and general size specifications*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1
access port
area of the *infusion container* (3.7) consisting of the *insertion point* (3.9) and the *injection point* (3.8), if applicable

© ISO 2026 – All rights reserved