International Standard

ISO 16053-2

First edition
2024-03

Paints and varnishes — Coating materials and coating systems for exterior wood —

Part 2: Exposure of wood coatings to artificial weathering using fluorescent UV lamps and water

Peintures et vernis — Produits de peinture et systèmes de peinture pour bois en extérieur —

Partie 2: Vieillissement artificiel des revêtements pour bois par exposition à des lampes UV fluorescentes et à de l'eau

Reference number
ISO 16053-2:2024(en)

© ISO 2024



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword ... **iv**
Introduction ... **vi**
1 Scope ... **1**
2 Normative references ... **1**
3 Terms and definitions ... **2**
4 Principle ... **2**
5 Apparatus ... **2**
5.1 Test chamber ... 2
5.2 Lamps ... 2
5.3 Device for wetting the test panels ... 2
5.4 Black panel thermometer ... 2
5.5 Irradiance control ... 3
6 Test panels ... **3**
6.1 Wood ... 3
6.2 Preparation and selection of wood panels ... 4
6.3 Preparation of coated panels ... 4
6.3.1 Wood conditioning ... 4
6.3.2 Preparation of panels for the test coating ... 4
6.3.3 Conditioning ... 5
7 Procedure ... **5**
7.1 Examination before exposure ... 5
7.2 Mounting the test panels ... 5
7.3 Exposure ... 5
7.3.1 Exposure cycle ... 5
7.3.2 Sample rotation and maintenance ... 6
7.3.3 Duration of test ... 6
7.4 Examination of test panels ... 6
8 Precision ... **6**
9 Expression of results and test report ... **9**
Annex A (normative) **Details of the test methods** ... **10**
Annex B (informative) **Explanatory notes** ... **12**
Annex C (normative) **Test for heartwood in pine** ... **13**
Annex D (informative) **Water purification** ... **14**
Annex E (normative) **Test for abnormally porous wood** ... **15**
Annex F (informative) **Alternative procedure for preparation and coating of panels** ... **16**
Annex G (informative) **Determination of adhesive strength of tape on test surface** ... **17**
Bibliography ... **18**

© ISO 2024 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by the European Committee for Standardization (CEN) (as EN 927-6:2018) and was adopted, without modification other than those given below. It was assigned to Technical Committee ISO/TC 35, *Paints and varnishes*, Subcommittee SC 9, *General test methods for paints and varnishes* and adopted under the "fast-track procedure".

— added CIE 1964 and CIE 1976 as normative references in Clause 2;

— removed redundant UVA-340 peak emission specification from 5.2;

— citation of Annex E added in 6.1;

— text below Figure 1 has been made into a Note;

— changed “mesh” to “grit” in 6.3.2;

— added a bibliography reference to the SERVOWOOD project in Clause 8;

— changed “may” to “can” in Clause 8 to indicate the possibility of the test precision to vary, rather than a permission;

— updated symbols and units and added a Note in Table 2;

— updated symbols in A.1;

— changed “guidance” to “instructions” in A.8.2.2;

— updated grammar in B.2 to improve clarity;

— removed hypothetical statement for testing other wood species from B.4;

— changed the status of Annex C from informative to normative;

— updated the title of Annex D;

— clarified the tape strength procedure in Annex G.

© ISO 2024 – All rights reserved

A list of all parts in the ISO 16053 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2024 – All rights reserved

Introduction

Coatings from paints, varnishes and similar materials are weathered in a laboratory in order to accelerate ageing processes (caused by temperature, wetness and irradiation) which occur during natural weathering. Generally, a simple accelerating ratio between ageing during artificial and natural weathering cannot be expected due to the influencing factors having different effects according to the nature of the coating and substrate. Predictable relationships can only be expected if the effect of the important parameters (spectral distribution of the irradiance in the photochemically relevant range, temperature of the specimen, type of wetting, wetting cycle relative humidity) on the coating is known. Moreover, acceleration of the coating chemistry can cause alternative degradation pathways to be followed. However, unlike natural weathering, testing in the laboratory can be controlled by the operator and therefore the results are more repeatable and reproducible. This document incorporates the results of a precision investigation that quantifies the capability of the test in terms of repeatability and reproducibility.

© ISO 2024 – All rights reserved

Paints and varnishes — Coating materials and coating systems for exterior wood —

Part 2: Exposure of wood coatings to artificial weathering using fluorescent UV lamps and water

1 Scope

This document specifies a method for determining the resistance of wood coatings to artificial weathering performed in an apparatus equipped with fluorescent UV lamps, condensation and water spray devices.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 554, *Standard atmospheres for conditioning and/or testing — Specifications*

ISO 2409, *Paints and varnishes — Cross-cut test*

ISO 2813, *Paints and varnishes — Determination of gloss value at 20°, 60° and 85°*

ISO 4618, *Paints and varnishes — Vocabulary*

ISO 4628-1:2016, *Paints and varnishes — Evaluation of degradation of coatings — Designation of quantity and size of defects, and of intensity of uniform changes in appearance — Part 1: General introduction and designation system*

ISO 4628-2, *Paints and varnishes — Evaluation of degradation of coatings — Designation of quantity and size of defects, and of intensity of uniform changes in appearance — Part 2: Assessment of degree of blistering*

ISO 4628-4, *Paints and varnishes — Evaluation of degradation of coatings — Designation of quantity and size of defects, and of intensity of uniform changes in appearance — Part 4: Assessment of degree of cracking*

ISO 4628-5, *Paints and varnishes — Evaluation of quantity and size of defects, and of intensity of uniform changes in appearance — Part 5: Assessment of degree of flaking*

ISO 4628-6, *Paints and varnishes — Evaluation of degradation of coatings — Designation of quantity and size of defects, and of intensity of uniform changes in appearance — Part 6: Assessment of degree of chalking by tape method*

ISO 16474-3, *Paints and varnishes — Methods of exposure to laboratory light sources — Part 3: Fluorescent UV lamps*

ISO 18314-1, *Analytical colorimetry — Part 1: Practical colour measurement*

EN 927-1, *Paints and varnishes — Coating materials and coating systems for exterior wood — Part 1: Classification and selection*

CIE 1964, *(U*, V*, W*) color space (CIEUVW)*

CIE 1976, *L*, u*, v* color space (CIELUV)*

© ISO 2024 – All rights reserved

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 4618 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

4 Principle

Artificial weathering of coatings using fluorescent UV lamps, condensation or water spray is carried out in order to produce a certain radiant exposure or mutually agreed total number of operation hours, based on a given degree of a change in a property or properties. The properties of the exposed coatings are compared with those of unexposed coatings, which are prepared from the same coating materials under identical conditions or with coatings whose degradation properties are known.

Radiation, temperature and humidity all contribute to the ageing process. Therefore, the apparatus specified in this document simulates all three factors.

The results obtained by this method do not necessarily directly relate to the results obtained under natural exposure conditions. The relationship between these results shall be established before the method can be used to predict performance. See B.3 for further explanations on correlation to natural weathering.

The standard test substrate is pine sapwood with the back side of panels coated. However, supplementary information on coating performance may be obtained by conducting optional tests on additional wood species, on pine, modified or impregnated by industrial processes or without coating the back side of the panels. See B.4 for further explanations on wood species.

5 Apparatus

5.1 Test chamber

The test chamber consists of an enclosure made from corrosion-resistant material which houses the lamps, a heated water tray, spray nozzles and test panel racks.

5.2 Lamps

An UV lamp emits UV radiation from a low-pressure mercury arc. The required spectral distribution is achieved by careful selection of the type of phosphor coating on the inner surface of the lamp and the nature of the glass used in the construction of the tubes.

NOTE The principal construction details are described in ISO 16474-1.

The lamp shall be of type UVA 340 in accordance with ISO 16474-3.

5.3 Device for wetting the test panels

The test panels shall be wetted by condensation from the heated water tray and by spray. To prevent spotting on to the test panels, water with a pH value between 5,0 and 7,5 and an electrical conductivity of maximum 2 mS/m, measured at (25 ± 1) °C shall be used. See Annex D.

5.4 Black panel thermometer

Set the apparatus to operate at the specified parameters. The temperature shall be monitored by a remote sensor attached to the black panel. The black panel thermometer shall be exposed to the same exposure

© ISO 2024 – All rights reserved