INTERNATIONAL STANDARD

ISO 17892-6

First edition
2017-02

Geotechnical investigation and testing — Laboratory testing of soil —

Part 6: Fall cone test

Reconnaissance et essais géotechniques — Essais de laboratoire sur les sols —

Partie 6: Essai de pénétration de cône

Reference number
ISO 17892-6:2017(E)

© ISO 2017



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword ..... **iv**
Introduction ..... **v**
1 Scope **1**
2 Normative references **1**
3 Terms and definitions **1**
4 Equipment **2**
5 Test procedure **4**
5.1 Test specimen preparation 4
5.2 Test points 5
5.3 Determination of fall cone penetration 5
6 Test results **7**
6.1 Average penetration 7
6.2 Estimated undrained fall cone shear strength 7
6.3 Additional estimate of undrained fall cone shear strength 7
7 Test report **7**
7.1 Mandatory reporting 7
7.2 Optional reporting 8
Annex A (normative) **Calibration, maintenance and checks** **9**
Bibliography **11**

© ISO 2017 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

ISO 17892-6 was prepared by the European Committee for Standardization (CEN) Technical Committee CEN/TC 341, *Geotechnical investigation and testing*, in collaboration with ISO Technical Committee ISO/TC 182, *Geotechnics*, in accordance with the agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This first edition cancels and replaces ISO/TS 17892-6:2004, which has been technically revised. It also incorporates the Technical Corrigendum ISO/TS 17892-6:2004/Cor 1:2006.

A list of all parts in the ISO 17892 series can be found on the ISO website.

© ISO 2017 – All rights reserved

Introduction

This document covers areas in the field of geotechnical engineering never previously standardized internationally. It is intended that this document presents broad good practice throughout the world and significant differences with national documents is not anticipated. It is based on international practice (see Reference [1]).

© ISO 2017 – All rights reserved

Geotechnical investigation and testing — Laboratory testing of soil —

Part 6:
Fall cone test

1 Scope

This document specifies a method of undrained strength index testing of both undisturbed and remoulded specimens of fine grained soils by the fall cone method.

This document is applicable to the laboratory estimation of undrained shear strength of a soil test specimen within the scope of geotechnical investigations.

In the fall cone test, a cone is allowed to fall with its tip towards a soil specimen, and the resulting penetration of the cone into the soil is measured. The penetration values are used to estimate the undrained shear strength. The fall cone test produces a complex shear in the test specimen, and does not represent either a vertical triaxial compression or a horizontal shear test. However, this index test may be correlated to some estimate of undrained shear strength determined in the laboratory by other test methods.

As the test is performed on a small laboratory specimen, the result may not agree with laboratory tests on larger specimens. In addition, the test specimen may not be fully representative of the soil in its natural state in the field; for example, the test specimen may not have fissures present *in situ* at a larger spacing than the specimen size.

Therefore, for the above reasons, the test can be regarded as an estimation of undrained shear strength, rather than a true measurement of it.

The ratio of the remoulded shear strength to the undisturbed shear strength may be used to estimate the sensitivity of a soil specimen. Time-dependent measurement of the shear strength may be used to assess the thixotropic regain of strength of a remoulded soil specimen.

NOTE This document fulfils the requirements of the strength index testing of soils for geotechnical investigation and testing in accordance with EN 1997–1 and EN 1997–2.

2 Normative references

The following documents are referred to in text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14688-1, *Geotechnical investigation and testing — Identification and classification of soil — Part 1: Identification and description*

ISO 17892-1, *Geotechnical investigation and testing — Laboratory testing of soil — Part 1: Determination of water content*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

© ISO 2017 – All rights reserved

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— IEC Electropedia: available at http://www.electropedia.org/

— ISO Online browsing platform: available at http://www.iso.org/obp

3.1
fall cone undrained shear strength of undisturbed soil
c_{ufc}
undrained shear strength of an undisturbed fine grained soil specimen estimated by the fall cone method

3.2
fall cone undrained shear strength of remoulded soil
c_{urfc}
undrained shear strength of a remoulded fine grained soil specimen estimated by the fall cone method

4 Equipment

See Annex A for calibration requirements for the equipment in this clause.

4.1 Cone apparatus

The apparatus shall permit the cone to be held firmly initially and to be released instantaneously to fall freely in the vertical direction into the soil specimen.

The apparatus shall have a mechanism which allows the cone to be raised or lowered and adjusted so that the tip of the cone just touches the surface of the specimen before the cone is released.

The cone apparatus shall be equipped with a method of measuring the penetration across the range of at least 4 mm to 20 mm penetration. Methods include a linear scale alongside the shaft, or a circular dial with a pointer needle or an electronic transducer, in both cases bearing on the upper end of the shaft. If the scale is designed for manual reading, it shall be marked in increments of 1 mm or better.

4.2 Fall cones

A set of cones with cone angles of 30° or 60° and masses (i.e. mass of cone plus shaft) covering a range of possible shear strengths shall be used. Typical examples of suitable fall cones are given on Table 1. There is limited experience or validation of cones of alternative configuration, for example, heavier than 400 g. Alternate cones should be used with caution.

Table 1 — Set of fall cones - typical total masses and tip angles

Mass	g	10	60	80	100	400
Tip angle β	°	60	60	30	30	30

A 60 g/60° cone is shown in Figure 1 as a typical example of such a cone.

© ISO 2017 – All rights reserved