INTERNATIONAL STANDARD

ISO 20569

First edition
2018-09

Dentistry — Trephine burs

Médecine bucco-dentaire — Fraises-trépans

Reference number
ISO 20569:2018(E)

© ISO 2018



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword ..... **iv**
Introduction ..... **v**
1 Scope **1**
2 Normative references **1**
3 Terms, definitions and symbols **1**
3.1 Terms and definitions 1
3.2 Symbols 2
4 Classifications **2**
5 Requirements **3**
5.1 Selection of metals 3
5.2 Surface finish 3
5.3 Dimensions 3
5.3.1 Number of teeth 3
5.3.2 Diameter of the working part and wall thickness 3
5.3.3 Length of the operative part 4
5.3.4 Overall length 4
5.3.5 Dimensions of the shank 4
5.4 Marking lines on the operative part 4
5.4.1 General 4
5.4.2 Thickness of scale line 4
5.5 Resistance to reprocessing 4
5.6 Vickers hardness 4
5.7 Run-out 4
5.8 Resistance to torque 5
6 Measurement and test methods **5**
6.1 Visual inspection 5
6.2 Dimensions and number of teeth 5
6.3 Resistance to reprocessing 5
6.4 Vickers hardness 6
6.5 Run-out 6
6.6 Resistance to torque 6
7 Marking, labelling and instructions for use **6**
7.1 Marking on the trephine bur 6
7.2 Labelling on the package 6
7.3 Instructions for use 6
Bibliography **7**

© ISO 2018 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

© ISO 2018 – All rights reserved

Introduction

Dental trephine burs are dental instruments used in dental implantology.

© ISO 2018 – All rights reserved

Dentistry — Trephine burs

1 Scope

This document specifies requirements and their test methods for trephine burs used in dentistry especially for oral implantology procedures such as collecting bone and/or removing an implant. It also specifies requirements for their marking and labelling.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1797, *Dentistry — Shanks for rotary and oscillating instruments*

ISO 1942, *Dentistry — Vocabulary*

ISO 6507-1, *Metallic materials — Vickers hardness test — Part 1: Test method*

ISO 7153-1, *Surgical instruments — Materials — Part 1: Metals*

ISO 8325:2004, *Dentistry — Test methods for rotary instruments*

ISO 13504, *Dentistry — General requirements for instruments and related accessories used in dental implant placement and treatment*

ISO 16443, *Dentistry — Vocabulary for dental implants systems and related procedure*

ISO 17664, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices*

3 Terms, definitions and symbols

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1797, ISO 1942, ISO 16443 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1.1
trephine bur
rotary instrument used together with a handpiece in oral implantology procedures such as preparing and collecting bone cores or removing implants

© ISO 2018 – All rights reserved

3.2 Symbols

D_1 outer diameter of working part

D_2 inner diameter of working part

D_3 diameter of shank

D_4 diameter of the operative part

L_1 overall length

L_2 length of shank

L_3 length of operative part

4 Classifications

For the purposes of this document, trephine burs shall be classified according to the purposes of use into the following types:

— Type 1: for bone collection (Figure 1);

— Type 2: for implant removal (Figure 2).



Key

1 working part with teeth

2 operative part

3 shank

4 marking lines

The design and number of lateral openings are left to the discretion of the manufacturer. They should be designed in such a way not to decrease the resistance nor the performance of the trephine bur.

Figure 1 — Trephine burs Type 1 for bone collection

© ISO 2018 – All rights reserved