INTERNATIONAL STANDARD

ISO/ASTM 51205

Third edition
2017-05

Practice for use of a ceric-cerous sulfate dosimetry system

Pratique de l'utilisation d'un système dosimétrique au sulfate cérique-céreux

Reference number
ISO/ASTM 51205:2017(E)

© ISO/ASTM International 2017



COPYRIGHT PROTECTED DOCUMENT

© ISO/ASTM International 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

ASTM International
100 Barr Harbor Drive, PO Box C700
West Conshohocken, PA 19428-2959, USA
Tel. +610 832 9634
Fax +610 832 9635
khooper@astm.org
www.astm.org

 © ISO/ASTM International 2017 – All rights reserved

Contents

Page

1 Scope **1**
2 Referenced documents **1**
3 Terminology **2**
4 Significance and use **2**
5 Effect of influence quantities **2**
6 Interferences **3**
7 Apparatus **3**
8 Reagents **3**
9 Preparation of the dosimeters **3**
10 Calibration of the dosimetry system **4**
11 Application of dosimetry system **6**
12 Minimum documentation requirements **6**
13 Measurement uncertainty **6**
14 Keywords **7**
Annexes **7**
Figure A1.1 Electrochemical cell **7**

© ISO/ASTM International 2017 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. International Standards are drafted in accordance with the editorial rules of ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ASTM International is one of the world's largest voluntary standards development organizations with global participation from affected stakeholders. ASTM technical committees follow rigorous due process balloting procedures.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and ASTM International shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by ASTM Committee E61 Radiation Processing and by Technical Committee ISO/TC 85, nuclear energy, nuclear technologies and radiological protection.

This third edition cancels and replaces the second edition (ISO/ASTM 51205:2009), which has been technically revised.

© ISO/ASTM International 2017 – All rights reserved

This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

ISO/ASTM 51205:2017(E)

Standard Practice for Use of a Ceric-Cerous Sulfate Dosimetry System[1]

This standard is issued under the fixed designation ISO/ASTM 51205; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

1. Scope

1.1 This practice covers the preparation, testing, and procedure for using the ceric-cerous sulfate dosimetry system to measure absorbed dose to water when exposed to ionizing radiation. The system consists of a dosimeter and appropriate analytical instrumentation. For simplicity, the system will be referred to as the ceric-cerous system. The ceric-cerous dosimeter is classified as a type 1 dosimeter on the basis of the effect of influence quantities. The ceric-cerous system may be used as a reference standard dosimetry system or as a routine dosimetry system.

1.2 This document is one of a set of standards that provides recommendations for properly implementing dosimetry in radiation processing, and describes a means of achieving compliance with the requirements of ISO/ASTM Practice 52628 for the ceric-cerous system. It is intended to be read in conjunction with ISO/ASTM Practice 52628.

1.3 This practice describes both the spectrophotometric and the potentiometric readout procedures for the ceric-cerous system.

1.4 This practice applies only to gamma radiation, X-radiation/bremsstrahlung, and high energy electrons.

1.5 This practice applies provided the following conditions are satisfied:

1.5.1 The absorbed-dose range is from 5×10^2 to 5×10^4 Gy **(1)**.[2]

1.5.2 The absorbed-dose rate does not exceed 10^6 Gy s^{-1} **(1)**.

1.5.3 For radionuclide gamma-ray sources, the initial photon energy is greater than 0.6 MeV. For bremsstrahlung photons, the initial energy of the electrons used to produce the bremsstrahlung photons is equal to or greater than 2 MeV. For electron beams, the initial electron energy is greater than 8 MeV.

NOTE 1—The lower energy limits are appropriate for a cylindrical dosimeter ampoule of 12-mm diameter. Corrections for dose gradient across the ampoule may be required for electron beams (**2**). The ceric-cerous system may be used at lower energies by employing thinner (in the beam direction) dosimeters (see ICRU Report 35).

1.5.4 The irradiation temperature of the dosimeter is above 0°C and below 62°C **(3)**.

NOTE 2—The temperature coefficient of dosimeter response is known only in this range (see 5.2). Use outside this range requires determination of the temperature coefficient.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

1.7 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced documents

2.1 *ASTM Standards:*[3]

C912 Practice for Designing a Process for Cleaning Technical Glasses

E170 Terminology Relating to Radiation Measurements and Dosimetry

E178 Practice for Dealing With Outlying Observations

E275 Practice for Describing and Measuring Performance of Ultraviolet and Visible Spectrophotometers

E666 Practice for Calculating Absorbed Dose From Gamma or X Radiation

E668 Practice for Application of Thermoluminescence-Dosimetry (TLD) Systems for Determining Absorbed Dose in Radiation-Hardness Testing of Electronic Devices

E925 Practice for Monitoring the Calibration of Ultraviolet-Visible Spectrophotometers whose Spectral Bandwidth does not Exceed 2 nm

E958 Practice for Estimation of the Spectral Bandwidth of

[1] This practice is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.02 on Dosimetry Systems, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved March 8, 2017. Published May 2017. Originally published as ASTM E1205–88. Last previous ASTM edition E1205–99. ASTM E1205–93 was adopted by ISO in 1998 with the intermediate designation ISO 15555:1998(E). The present International Standard ISO/ASTM 51205:2017(E) is a major revision of ISO/ASTM 51205-2009(E).

[2] The boldface numbers in parentheses refer to the bibliography at the end of this standard.

[3] For referenced ASTM and ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

© ISO/ASTM International 2017 – All rights reserved

Ultraviolet-Visible Spectrophotometers

2.2 *ISO/ASTM Standards:*[3]

51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

51707 Guide for Estimation of Measurement Uncertainty in Dosimetry for Radiation Processing

52628 Practice for Dosimetry in Radiation Processing

52701 Guide for Performance Characterization of Dosimeters and Dosimetry Systems for Use in Radiation Processing

2.3 *ISO Standards:*[4]

12749-4 Nuclear energy – Vocabulary – Part 4: Dosimetry for radiation processing

2.4 *ISO/IEC Standards:*[4]

17025 General Requirements for the Competence of Testing and Calibration Laboratories

2.5 *Joint Committee for Guides in Metrology (JCGM) Reports:*

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement[5]

JCGM 200:2012 (JCGM 200:2008 with minor revisions), VIM, International Vocabulary of Metrology – Basis and General Concepts and Associated Terms[6]

2.6 *International Commission on Radiation Units and Measurements (ICRU) Reports:*[7]

ICRU Report 10b (NBS Handbook 85) Physical Aspects of Irradiation

ICRU Report 35 Radiation Dosimetry: Electron Beams with Initial Energies Between 1 and 50 MeV

ICRU Report 80 Dosimetry Systems for Use in Radiation Processing

ICRU Report 85a Fundamental Quantities and Units for Ionizing Radiation

3. Terminology

3.1 *Definitions:*

3.1.1 *approved laboratory*—laboratory that is a recognized national metrology institute, or has been formally accredited to ISO/IEC 17025, or has a quality system consistent with the requirements of ISO/IEC 17025.

3.1.1.1 *Discussion*—A recognized national metrology institute or other calibration laboratory accredited to ISO/IEC 17025 should be used in order to ensure traceability to a national or international standard. A calibration certificate provided by a laboratory not having formal recognition or accreditation will not necessarily be proof of traceability to a national or international standard.

3.1.2 *ceric-cerous dosimeter*—specially prepared solution of ceric sulfate and cerous sulfate in sulfuric acid, individually sealed in an appropriate container such as a glass ampoule, where the radiation-induced changes in electropotential or optical absorbance of the solution are related to absorbed dose to water.

3.1.3 *molar linear absorption coefficient*, ε_m—constant relating the spectrophotometric absorbance, A_λ, of an optically absorbing molecular species at a given wavelength, λ, per unit pathlength, d, to the molar concentration, c, of that species in solution:

$$\varepsilon_m = \frac{A_\lambda}{d \cdot c} \tag{1}$$

SI unit: $m^2\ mol^{-1}$

3.1.3.1 *Discussion*—The measurement is sometimes expressed in units of $L\ mol^{-1}\ cm^{-1}$.

3.1.4 *radiation chemical yield, G(x)*—quotient of $n(x)$ by $\bar{\varepsilon}$, where $n(x)$ is the mean amount of a specified entity, x, produced, destroyed, or changed by the mean energy, $\bar{\varepsilon}$, imparted to the matter.

$$G(x) = \frac{n(x)}{\bar{\varepsilon}} \tag{2}$$

SI unit: $mol\ J^{-1}$

3.1.5 *reference standard dosimetry system*—dosimetry system, generally having the highest metrological quality available at a given location or in a given organization, from which measurements made there are derived.

3.1.6 *type 1 dosimeter*—dosimeter of high metrological quality, the response of which is affected by individual influence quantities in a well-defined way that can be expressed in terms of independent correction factors.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *electropotential, E*—difference in potential between the solutions in the two compartments of an electrochemical cell, measured in millivolts.

3.3 Definitions of other terms used in this practice that pertain to radiation measurement and dosimetry may be found in ISO 12749-4, ASTM Terminology E170, ICRU 85a, and VIM; these documents, therefore, may be used as alternative references.

4. Significance and use

4.1 The ceric-cerous system provides a reliable means for determining absorbed dose to water. It is based on a process of reduction of ceric ions to cerous ions in acidic aqueous solution by ionizing radiation **(1, 4, ICRU Report 80)**.

NOTE 3—The ceric-cerous system described in the practice has cerous sulfate added to the initial solution to reduce the effect of organic impurities and to allow the potentiometric method of measurement. Other systems used for dosimetry include solutions of ceric sulfate or ceric ammonium sulfate in sulfuric acid without the initial addition of cerous sulfate. These other systems are based on the same process of reduction of ceric ions to cerous ions but are not included in this practice.

5. Effect of influence quantities

5.1 Guidance on the determination of the performance characteristics of dosimeters and dosimetry systems can be

[4] Available from International Organization for Standardization (ISO), ISO Central Secretariat, BIBC II, Chemin de Blandonnet 8, CP 401, 1214 Vernier, Geneva, Switzerland, http://www.iso.org.

[5] Document produced by Working Group 1 of the Joint Committee for Guides in Metrology (JCGM WG1), Available free of charge at the BIPM website (http://www.bipm.org).

[6] Document produced by Working Group 2 of the Joint Committee for Guides in Metrology (JCGM WG2), Available free of charge at the BIPM website (http://www.bipm.org).

[7] Available from International Commission on Radiation Units and Measurements, 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814, USA.

© ISO/ASTM International 2017 – All rights reserved