INTERNATIONAL STANDARD

ISO/ASTM 51940

Fourth edition
2022-08

Guidance for dosimetry for sterile insects release programs

Lignes directrices de la dosimétrie pour des programmes de lâchers d'insectes stériles

Reference number
ISO/ASTM 51940:2022(E)

© ISO/ASTM International 2022



COPYRIGHT PROTECTED DOCUMENT

© ISO/ASTM International 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11

Email: copyright@iso.org
Website: www.iso.org

ASTM International
100 Barr Harbor Drive, PO Box C700
West Conshohocken, PA 19428-2959, USA
Phone: +610 832 9634
Fax: +610 832 9635
Email: khooper@astm.org
Website: www.astm.org

Published in Switzerland

© ISO/ASTM International 2022 – All rights reserved

Contents

Page

1 Scope **1**
2 Referenced documents **2**
3 Terminology **2**
4 Significance and use **4**
5 Types of facilities and modes of operation **4**
6 Radiation source characteristics **5**
7 Dosimetry systems **6**
8 Installation and operational qualification **7**
9 Performance qualification **8**
10 Routine product processing **9**
11 Measurement uncertainty **10**
12 Keywords **10**
Annex **11**
Table 1 Examples of reference-standard dosimetry systems **6**
Table 2 Examples of routine dosimetry systems **7**
Table A1.1 Recommended Procedures **11**

© ISO/ASTM International 2022 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. International Standards are drafted in accordance with the editorial rules of ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ASTM International is one of the world's largest voluntary standards development organizations with global participation from affected stakeholders. ASTM technical committees follow rigorous due process balloting procedures.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO and ASTM International shall be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by ASTM Committee E61 Radiation Processing and by Technical Committee ISO/TC 85, nuclear energy, nuclear technologies and radiological protection.

This fourth edition cancels and replaces the third edition (ISO/ASTM 51940:2013), which has been technically revised.

© ISO/ASTM International 2022 – All rights reserved

This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

ISO/ASTM 51940:2022(E)

Standard Guidance for Dosimetry for Sterile Insect Release Programs[1]

This standard is issued under the fixed designation ISO/ASTM 51940; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

INTRODUCTION

The purpose of this document is to present information on the use of ionizing energy (radiation) for the radiation-induced reproductive sterilization of live insects for use in pest management programs.

This document is intended to serve as a recommendation to be followed when using irradiation technology where approved by an appropriate regulatory authority. It is not to be construed as a requirement for the use of irradiation nor as a required code of practice. While the use of irradiation involves certain essential requirements to attain the objective of the treatment, some parameters can be varied in optimizing the process.

1. Scope

1.1 This document outlines dosimetric procedures to be followed for the radiation-induced reproductive sterilization of live insects for use in pest management programs. The primary use of such insects is in the Sterile Insect Technique, where large numbers of reproductively sterile insects are released into the field to mate with and thus control pest populations of the same species. A secondary use of sterile insects is as benign hosts for rearing insect parasitoids. A third use is for testing detection traps for fruit flies and moths, and testing mating disruption products for moths. The procedures outlined in this document will help ensure that insects processed with ionizing radiation from gamma, electron, or X-ray sources receive absorbed doses within a predetermined range. Information on effective dose ranges for specific applications of insect sterilization, or on methodology for determining effective dose ranges, is not within the scope of this document.

NOTE 1—Dosimetry is only one component of a total quality assurance program to ensure that irradiated insects are adequately sterilized and fully competitive or otherwise suitable for their intended purpose.

1.2 This document provides information on dosimetry for the irradiation of insects for these types of irradiators: self-contained dry-storage ^{137}Cs or ^{60}Co irradiators, self-contained low-energy X-ray irradiators (maximum processing energies from 150 keV to 300 keV), large-scale gamma irradiators, and electron accelerators (electron and X-ray modes).

NOTE 2—Additional, detailed information on dosimetric procedures to be followed in installation qualification, operational qualification, performance qualification, and routine product processing can be found in ISO/ASTM Practices 51608 (X-ray [bremsstrahlung] facilities processing at energies over 300 keV), 51649 (electron beam facilities), 51702 (large-scale gamma facilities), and 52116 (self-contained dry-storage gamma facilities), and in Ref (**1**)[2] (self-contained X-ray facilities).

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard except for the non-SI units of minute (min) hour (h) and day (d). These non-SI units are accepted for use within the SI system.

1.4 This document is one of a set of standards that provides recommendations for properly implementing and utilizing radiation processing. It is intended to be read in conjunction with ISO/ASTM Practice 52628.

1.5 The absorbed dose for insect sterilization is typically within the range of 20 Gy to 600 Gy.

1.6 This document refers, throughout the text, specifically to reproductive sterilization of insects. It is equally applicable to radiation sterilization of invertebrates from other taxa (for example, Acarina, Gastropoda) and to irradiation of live insects or other invertebrates for other purposes (for example, inducing mutations), provided the absorbed dose is within the range specified in 1.5.

1.7 This document also covers the use of radiation-sensitive indicators for the visual and qualitative indication that the insects have been irradiated (see ISO/ASTM Guide 51539).

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

[1] This document is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.04 on Specialty Application, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved May 20, 2022. Published August 2022. Originally published as ASTM E 1940–98. The present International Standard ISO/ASTM 51940:2022(E) replaces and is a major revision of the last previous edition ISO/ASTM 51940:2013(E).

[2] The boldface numbers in parentheses refer to the list of references at the end of this standard.

© ISO/ASTM International 2022 – All rights reserved

1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced documents

2.1 *ASTM Standards:*[3]

E3083 Terminology Relating to Radiation Processing: Dosimetry and Applications

2.2 *ISO/ASTM Standards:*[3]

51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

51275 Practice for Use of a Radiochromic Film Dosimetry System

51310 Practice for Use of a Radiochromic Optical Waveguide Dosimetry System

51539 Guide for Use of Radiation-Sensitive Indicators

51607 Practice for Use of an Alanine-EPR Dosimetry System

51608 Practice for Dosimetry in an X-Ray (Bremsstrahlung) Facility for Radiation Processing at Energies Between 50 keV and 7.5 MeV

51649 Practice for Dosimetry in an Electron Beam Facility for Radiation Processing at Energies Between 300 keV and 25 MeV

51702 Practice for Dosimetry in a Gamma Facility for Radiation Processing

51707 Guide for Estimation of Measurement Uncertainty in Dosimetry for Radiation Processing

51956 Practice for Use of a Thermoluminescence-Dosimetry System (TLD System) for Radiation Processing

52116 Practice for Dosimetry for a Self-Contained Dry-Storage Gamma-Ray Irradiator

52303 Guide for Absorbed-Dose Mapping in Radiation Processing Facilities

52628 Practice for Dosimetry in Radiation Processing

52701 Guide for Performance Characterization of Dosimeters and Dosimetry Systems for Use in Radiation Processing

2.3 *International Commission on Radiation Units and Measurements (ICRU) Reports:*[4]

ICRU Report 80 Dosimetry Systems for Use in Radiation Processing

ICRU 85a Fundamental Units and Quantities for Ionizing Radiation

2.4 *ISO Standards:*[5]

ISO/IEC 17025 General Requirements for the Competence of Testing and Calibration Laboratories

ISO 12749-4 Nuclear energy – Vocabulary – Part 4: Dosimetry for radiation processing

2.5 *Joint Committee for Guides in Metrology (JCGM) Reports:*

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement[6]

JCGM 200:2012, VIM International Vocabulary of Metrology – Basic and General Concepts and Associated Terms[7]

3. Terminology

3.1 *Definitions:*

3.1.1 *absorbed dose (D)*—quotient of $d\bar{\varepsilon}$ by dm, where $d\bar{\varepsilon}$ is the mean energy imparted by ionizing radiation to matter of mass dm thus

$$D = d\bar{\varepsilon}/dm$$

3.1.1.1 *Discussion*—The SI unit of absorbed dose is the gray (Gy), where 1 gray is equivalent to the absorption of 1 joule per kilogram of the specified material (1 Gy = 1 J / kg).

3.1.2 *absorbed-dose mapping*—measurement of absorbed-dose within an irradiated product to produce a one-, two- or three-dimensional distribution of absorbed dose, thus rendering a map of absorbed-dose values.

3.1.3 *absorbed-dose rate, $\dot{D}$*—absorbed dose in a material per incremental time interval, that is, the quotient of dD by dt. Also see ASTM Terminology E3083. The SI unit is $Gy \cdot s^{-1}$

$$\dot{D} = dD/dt$$

3.1.3.1 *Discussion*—The absorbed-dose rate can be specified in terms of its average value over long-time intervals, for example in units of $Gy \cdot min^{-1}$ or $Gy \cdot h^{-1}$

3.1.4 *approved laboratory*—laboratory that is a recognized national metrology institute, or has been formally accredited to ISO/IEC 17025, or has a quality system consistent with the requirements of ISO/IEC 17025.

3.1.4.1 *Discussion*—A recognized national metrology institute or other calibration laboratory accredited to ISO/IEC 17025 should be used in order to ensure traceability to a national or international standard. A calibration certificate provided by a laboratory not having formal recognition or accreditation will not necessarily be proof of traceability to a national or international standard.

3.1.5 *calibration [VIM, 6.11]*—set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring instrument or measuring system, or values represented by a material measure or a reference material, and the corresponding values realized by standards.

3.1.5.1 *Discussion*—Calibration conditions include environmental and irradiation conditions present during irradiation, storage and measurement of the dosimeters that are used for the

[3] For referenced ASTM and ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

[4] Available from the International Commission on Radiation Units and Measurements, 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814, USA.

[5] Available from International Organization for Standardization (ISO), 1 Rue de Varembé, Case Postale 56, CH-1211, Geneva 20, Switzerland.

[6] Document produced by Working Group 1 of the Joint Committee for Metrology (JCGM/WG 1). Available free of charge at the BIPM website (http://www.bipm.org).

[7] Document produced by Working Group 2 of the Joint Committee for Metrology (JCGM/WG 2). Available free of charge at the BIPM website (http://www.bipm.org).

© ISO/ASTM International 2022 – All rights reserved