FINAL
DRAFT

INTERNATIONAL STANDARD IEC/FDIS 80601-2-60

ISO/TC **106**/SC **6**

Secretariat: DIN

Voting begins on:
2019-02-01

Voting terminates on:
2019-03-15

Medical electrical equipment —

Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment

Appareils électromédicaux —

Partie 2-60: Exigences particulières pour la sécurité de base et les performances essentielles des équipements dentaires

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

This draft is submitted to a parallel vote in ISO and in IEC.

Reference number
IEC/FDIS 80601-2-60:2019(E)

© IEC 2019

CONTENTS

FOREWORD 3
201.1 Scope, object and related standards 6
201.2 Normative references 7
201.3 Terms and definitions 8
201.4 General requirements 9
201.5 General requirements for testing of ME EQUIPMENT 9
201.6 Classification of ME EQUIPMENT and ME SYSTEMS 10
201.7 ME EQUIPMENT identification, marking and documents 10
201.8 Protection against electrical HAZARDS from ME EQUIPMENT 11
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS 16
201.10 Protection against unwanted and excessive radiation HAZARDS 18
201.11 Protection against excessive temperatures and other HAZARDS 18
201.12 Accuracy of controls and instruments and protection against hazardous outputs 22
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT 22
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) 23
201.15 Construction of ME EQUIPMENT 23
201.16 ME SYSTEMS 23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS 23
201.101 Cordless HAND-HELD and foot-operated control devices 24
Annexes 25
Annex AA (informative) Particular guidance and rationale 26
Bibliography 38
Index of defined terms used in this particular standard 39

Figure AA.1 – Example of APPLIED PARTS for DENTAL EQUIPMENT 27
Figure AA.2 – Calculation of LEAKAGE CURRENT 28
Figure AA.3 – Insulation problem of commutator DENTAL ELECTRICAL MOTOR 30
Figure AA.4 – Loading fan construction 36
Figure AA.5 – Load diagram with loading fan 36

Table 201.101 – Test voltages for solid insulation for SECONDARY CIRCUITS according to 201.8.9.1.12 12
Table 201.102 – Determination of TENSILE SAFETY FACTOR 17
Table 201.103 – Mass distribution 18
Table 201.104 – Allowable maximum temperatures for DENTAL HANDPIECES 19
Table AA.1 – RATED impulse voltage for equipment energized directly from the low-voltage mains 31

INTERNATIONAL ELECTROTECHNICAL COMMISSION

_____

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-60 has been prepared by a Joint Working Group of subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and subcommittee 6: Dental equipment, of ISO technical committee 106: Dentistry.

This second edition cancels and replaces the first edition published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

a) alignment with IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
62D/XX/FDIS	62D/XX/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

– requirements and definitions: roman type;
– *test specifications: italic type*;
– informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
– TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

– “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
– “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

– “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
– “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
– “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment

201.1 Scope, object and related standards

Clause 1 of the general standard[1] applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 80601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE OF DENTAL UNITS, DENTAL PATIENT CHAIRS, DENTAL HANDPIECES AND DENTAL OPERATING LIGHTS, hereafter referred to as DENTAL EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for DENTAL EQUIPMENT (as defined in 201.3.202.)

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-3 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular

---

1 The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.*

ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"*Replacement*" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"*Addition*" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"*Amendment*" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows*:*

Replacement:

IEC 60825-1:2014, *Safety of laser products – Part 1: Equipment classification and requirements*

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-2:2017, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

IEC 60601-2-22:2007, *Medical electrical equipment – Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment*

IEC 60601-2-57:2011, *Medical electrical equipment – Part 2-57: Particular requirements for the basic safety and essential performance of non-laser light source equipment intended for therapeutic, diagnostic, monitoring and cosmetic/aesthetic use*

IEC 60664-1:2007, *Insulation coordination for equipment within low-voltage systems – Part 1: Principles, requirements and tests*

IEC 60664-4:2005, *Insulation coordination for equipment within low-voltage systems – Part 4: Consideration of high-frequency voltage stress*

IEC 61180:2016, *High-voltage test techniques for low-voltage equipment – Definitions, test and procedure requirements, test equipment*

IEC 61810-1:2015, *Electromechanical elementary relays – Part 1: General and safety requirements*

ISO 1942:2009, *Dentistry – Vocabulary*

ISO 14457:2012, *Dentistry – Handpieces and motors*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, IEC 60601-2-2:2017 and ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at http://www.electropedia.org/
- ISO Online browsing platform: available at http://www.iso.org/obp

NOTE An index of defined terms is found beginning on page 39.

Addition:

201.3.201
DENTAL ELECTRICAL MOTOR
HAND-HELD part of the DENTAL HANDPIECE electrically powered by the DENTAL UNIT

201.3.202
DENTAL EQUIPMENT
ME EQUIPMENT with any combination of DENTAL HANDPIECES, DENTAL UNITS, DENTAL PATIENT CHAIRS and DENTAL OPERATING LIGHTS

201.3.203
DENTAL HANDPIECE
HAND-HELD instrument used in dentistry for use in PATIENT treatment and connected to the DENTAL UNIT

201.3.204
DENTAL OPERATING LIGHT
device designed for use by an OPERATOR for illuminating the oral cavity, consisting of a luminaire and one or more lamps

201.3.205
DENTAL PATIENT CHAIR
device designed to support and position the PATIENT for treatment and therefore provided with a range of movements

201.3.206
DENTAL UNIT
device (stationary or non-stationary) through which electrical power and/or various fluids or gasses are supplied to at least one or more DENTAL HANDPIECE(S) and devices

Note 1 to entry: This device is usually fitted with conveniently oriented instrument holders and controls, and consists of interconnected sub-units of DENTAL EQUIPMENT and instruments providing a functional unit for dental use.

201.3.207
OPERATOR SIDE OF DENTAL HANDPIECE
part of DENTAL HANDPIECE which is designed to be handheld by the OPERATOR in NORMAL USE

201.3.208
PATIENT SIDE OF DENTAL HANDPIECE
part of DENTAL HANDPIECE which is designed to be introduced into the oral cavity where all parts of the DENTAL HANDPIECE within 80 mm of the tip shall be considered as an APPLIED PART

201.3.209
CORD-CONNECTED MOBILE PARTS OF DENTAL EQUIPMENT
DENTAL EQUIPMENT which is permanently connected to FIXED parts of DENTAL EQUIPMENT and is equipped with castor or wheels for positioning by the OPERATOR and which is intended for the use on an even, obstacle free, floor

201.4 General requirements

Clause 4 of the general standard applies, except as follows:

201.4.3 ESSENTIAL PERFORMANCE

Replacement:

DENTAL EQUIPMENT does not have ESSENTIAL PERFORMANCE.

201.5 General requirements for testing of ME EQUIPMENT

Clause 5 of the general standard applies, except as follows:

201.5.3 Ambient temperature, humidity, atmospheric pressure

Amendment of item a):