FINAL
DRAFT

INTERNATIONAL STANDARD

IEC/FDIS 80369-5

ISO/TC **210**

Secretariat: **ANSI**

Voting begins on:
2015-12-11

Voting terminates on:
2016-02-11

Small-bore connectors for liquids and gases in healthcare applications —

Part 5:
Connectors for limb cuff inflation applications

Raccords de petite taille pour liquides et gaz utilisés dans le domaine de la santé —

Partie 5: Raccords destinés à des applications au gonflement de ballonnet

Please see the administrative notes on page ii

RECIPIENTS OF THIS DRAFT ARE INVITED TO SUBMIT, WITH THEIR COMMENTS, NOTIFICATION OF ANY RELEVANT PATENT RIGHTS OF WHICH THEY ARE AWARE AND TO PROVIDE SUPPORTING DOCUMENTATION.

IN ADDITION TO THEIR EVALUATION AS BEING ACCEPTABLE FOR INDUSTRIAL, TECHNOLOGICAL, COMMERCIAL AND USER PURPOSES, DRAFT INTERNATIONAL STANDARDS MAY ON OCCASION HAVE TO BE CONSIDERED IN THE LIGHT OF THEIR POTENTIAL TO BECOME STANDARDS TO WHICH REFERENCE MAY BE MADE IN NATIONAL REGULATIONS.

Reference number
IEC/FDIS 80369-5:2015(E)

This final draft is submitted to a parallel approval vote in ISO and IEC. Each ISO member body and IEC national committee is requested to take appropriate steps to harmonize the national viewpoint in order to cast the same “yes” or “no” vote to both ISO and IEC.

Positive votes shall not be accompanied by comments.

Negative votes shall be accompanied by the relevant technical reasons.

© International Electrotechnical Commission

CONTENTS

FOREWORD 4
INTRODUCTION 6
1 * Scope 7
2 Normative references 7
3 Terms and definitions 8
4 General requirements 9
4.1 General requirements for the limb cuff inflation APPLICATION 9
4.2 Material used for SMALL-BORE CONNECTORS 10
4.3 TYPE TESTS 10
5 Dimensional requirements for sphygmomanometer and cuff SMALL-BORE CONNECTORS 10
5.1 * Requirements for adult or paediatric PATIENT SMALL-BORE CONNECTORS (S1) 10
5.2 Void 11
6 Performance requirements 11
6.1 Air leakage 11
6.2 * Resistance to separation from axial load 11
Annex A (informative) Rationale and guidance 12
A.1 General guidance 12
A.2 Rationale for particular clauses and subclauses 12
Annex B (normative) SMALL-BORE CONNECTORS for the limb cuff inflation APPLICATION 14
Annex C (normative) Reference CONNECTORS 18
C.1 General requirements for reference CONNECTORS 18
C.2 * Sphygmomanometer and cuff S1 reference CONNECTORS 18
Annex D (informative) Assessment of MEDICAL DEVICES and their attributes with CONNECTIONS within this APPLICATION 19
Annex E (informative) Summary of the usability requirements for SMALL-BORE CONNECTORS for limb cuff inflation APPLICATIONS 20
E.1 USER PROFILE 20
E.2 Use scenarios 21
E.3 Use environments and scenarios 21
E.4 Generic USER needs 21
Annex F (informative) Summary of SMALL-BORE CONNECTOR design requirements for limb cuff inflation APPLICATIONS 22
Annex G (informative) Summary of assessment of the design of the CONNECTORS for limb cuff inflation APPLICATION 30
G.1 General 30
G.2 Summary of the engineering analysis of the design 30
G.2.1 NON-INTERCONNECTABLE analysis 30
G.2.2 S1 male to N1 male 30
G.3 Summary of the design VERIFICATION 30
G.4 Summary of the design validation 31
G.5 Summary of the design review 31
Annex H (informative) Obsolete limb cuff inflation CONNECTOR 32
Annex I (informative) Air leakage by pressure decay TEST METHOD 35
I.1 Principle 35

I.2 * Test conditions 35
I.2.1 Test sample preconditioning 35
I.2.2 Environmental test conditions 35
I.3 Apparatus 35
I.4 PROCEDURE 35
I.5 Test report 36
Annex J (informative) Resistance to separation from axial load TEST METHOD 38
J.1 Principle 38
J.2 * Test conditions 38
J.2.1 Test sample preconditioning 38
J.2.2 Environmental test conditions 38
J.3 Apparatus 38
J.4 PROCEDURE 38
J.5 Test report 38
Annex K (informative) Reference to the essential principles 40
Index of defined terms 42
Bibliography 43

Figure B.1 – Male cuff S1 SMALL-BORE CONNECTOR 14
Figure B.2 – Female sphygmomanometer S1 SMALL-BORE CONNECTOR 16
Figure B.3 – Sphygmomanometer and cuff SMALL-BORE CONNECTOR (S1) assembly 16
Figure H.1 – Obsolete sphygmomanometer and cuff SMALL-BORE CONNECTOR 33

Table B.1 – Male cuff S1 SMALL-BORE CONNECTOR dimensions 15
Table B.2 – Female sphygmomanometer S1 SMALL-BORE CONNECTOR dimensions 17
Table E.1 – USER PROFILE 20
Table F.1 – Adult or paediatric PATIENT sphygmomanometer and cuff S1 CONNECTOR-specific design requirements 22
Table F.2 – Neonatal sphygmomanometer and cuff CONNECTOR-specific design requirements 26
Table G.1 – Summary of possible misconnection from CAD analysis 30
Table H.1 – Obsolete male sphygmomanometer and cuff SMALL-BORE CONNECTOR dimensions 34
Table H.2 – Obsolete female sphygmomanometer and cuff SMALL-BORE CONNECTOR dimensions 34
Table K.1 – Correspondence between this document and the essential principles 40

INTERNATIONAL ELECTROTECHNICAL COMMISSION

_____

SMALL-BORE CONNECTORS FOR LIQUIDS AND GASES IN HEALTHCARE APPLICATIONS –

Part 5: Connectors for limb cuff inflation applications

FOREWORD

1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
6) All users should ensure that they have the latest edition of this publication.
7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC/ISO 80369-5 has been prepared by a Joint Working Group of subcommittee 62D: Electromedical equipment of IEC technical committee 62: Electrical equipment in medical practice, ISO technical committee 210, Quality management and corresponding general aspects for medical devices and CEN/CENELEC TC3/WG 2, Small-bore connectors.

The text of this standard is based on the following documents:

FDIS	Report on voting
XX/XX/FDIS	XX/XX/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC/ISO 80369 series, published under the general title *Small-bore connectors for liquids and gases in healthcare applications*, can be found on the IEC and ISO websites.

In this standard, the following print types are used:

– Requirements and definitions: roman type.

– Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

– TERMS DEFINED IN CLAUSE 3 OF THIS STANDARD OR AS NOTED IN THE INDEX OF DEFINED TERMS: SMALL CAPITALS.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

– “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;

– “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;

– “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

NOTE The attention of Member Bodies and National Committees is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised ISO or IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

The National Committees are requested to note that for this publication the stability date is 2019

THIS TEXT IS INCLUDED FOR THE INFORMATION OF THE NATIONAL COMMITTEES AND WILL BE DELETED AT THE PUBLICATION STAGE.

INTRODUCTION

This International Standard was developed because of several incidents, with catastrophic consequences, resultant from inappropriate medication, liquid nutritional formula or air being administered intravenously. Many incidents have been reported leading to international recognition of the importance of these issues, and a need has been identified to develop specific CONNECTORS for MEDICAL DEVICES and their ACCESSORIES used to deliver fluids in other APPLICATIONS.

The IEC/ISO 80369 series was developed to prevent misconnection between SMALL-BORE CONNECTORS used in different APPLICATIONS. Part 1 specifies the requirements necessary to verify the designs and dimensions of SMALL-BORE CONNECTORS to ensure that:

a) they do not misconnect with other SMALL-BORE CONNECTORS; and

b) they safely and securely connect with their mating half.

Part 20 contains the common TEST METHODS to support the performance requirements for SMALL-BORE CONNECTORS. The other parts specify the designs of SMALL-BORE CONNECTORS for the various APPLICATIONS.

This part of ISO 80369 specifies the design and the dimensions and drawings of SMALL-BORE CONNECTORS intended for use in limb cuff inflation APPLICATIONS. The informative Annex D through Annex G describe the methods by which this design has been assessed. Other parts of ISO 80369 include requirements for SMALL-BORE CONNECTORS used in different APPLICATION categories.

CONNECTORS manufactured to the dimensions set out within this International Standard are therefore dimensionally incompatible with the SMALL-BORE CONNECTORS used in other APPLICATIONS specified by the standards in this series, unless otherwise indicated. If fitted to the relevant MEDICAL DEVICES and ACCESSORIES, these CONNECTORS should be able to prevent air being delivered intravenously. CONNECTORS manufactured to the dimensions specified in this standard are also NON-INTERCONNECTABLE with any of the other CONNECTORS identified in the ISO 80369 series of standards for SMALL-BORE CONNECTORS, unless otherwise indicated.

SMALL-BORE CONNECTORS FOR LIQUIDS AND GASES IN HEALTHCARE APPLICATIONS –

Part 5: Connectors for limb cuff inflation applications

1 * Scope

This part of International Standard 80369 specifies dimensions and requirements for the design and functional performance of SMALL-BORE CONNECTORS intended to be used for CONNECTIONS in limb cuff inflation APPLICATIONS of MEDICAL DEVICES and ACCESSORIES. Limb cuff inflation APPLICATIONS include CONNECTIONS between a sphygmomanometer and its cuff. [3] [7] [1]

This part of International Standard 80369 does not specify requirements for the MEDICAL DEVICES or ACCESSORIES that use these CONNECTORS. Such requirements are given in particular International Standards for specific MEDICAL DEVICES or ACCESSORIES.

This part of International Standard 80369 does not specify requirements for pressurizing and depressurizing the retention mechanism (e.g. balloon) used to hold invasive MEDICAL DEVICES in place.

NOTE 1 MANUFACTURERS are encouraged to incorporate the SMALL-BORE CONNECTORS specified in this part of ISO 80369 into MEDICAL DEVICES, medical systems or ACCESSORIES, even if currently not required by the relevant particular MEDICAL DEVICE standards. It is expected that when the relevant particular MEDICAL DEVICE standards are revised, requirements for SMALL-BORE CONNECTORS, as specified in this part of ISO 80369 will be included.

NOTE 2 The requirements for SMALL-BORE CONNECTORS intended to be used with neonatal PATIENTS to connect a cuff to a sphygmomanometer are intended to be added to this standard by an amendment or new edition. IEC 80601-2-30 [7] defines the age range for neonatal mode usage of sphygmomanometers.

NOTE 3 The requirements for SMALL-BORE CONNECTORS intended to be used to connect a tourniquet to its inflating equipment are intended to be added to this standard by an amendment or new edition.

NOTE 4 MANUFACTURERS are encouraged to incorporate the SMALL-BORE CONNECTORS specified in this part of ISO 80369 into MEDICAL DEVICES or ACCESSORIES, even if currently not required by the relevant particular MEDICAL DEVICE standards. It is expected that when the relevant particular MEDICAL DEVICE standards are revised, requirements for SMALL-BORE CONNECTORS, as specified in ISO 80369, will be included.

NOTE 5 ISO 80369-1:2010, 5.8, specifies alternative methods of compliance with ISO 80369-1:2010, for SMALL-BORE CONNECTORS intended for in limb cuff inflation APPLICATIONS of MEDICAL DEVICES or ACCESSORIES, which do not comply with this part of ISO 80369.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the bibliography on page 43.

ISO 5356-1:2004, *Anaesthetic and respiratory equipment – Conical connectors – Part 1: Cones and sockets*

---

1 Figures in square brackets refer to the Bibliography.

ISO 5356-1:2015[2], *Anaesthetic and respiratory equipment – Conical connectors – Part 1: Cones and sockets*

ISO 5356 2:2006[3], *Anaesthetic and respiratory equipment – Conical connectors – Part 2: Screw-threaded weight-bearing connectors*

ISO 5356-2:2012, *Anaesthetic and respiratory equipment – Conical connectors – Part 2: Screw-threaded weight-bearing connectors*

ISO 8185:2007, *Respiratory tract humidifiers for medical use – Particular requirements for respiratory humidification systems*

ISO 14971:2007, *Medical devices – Application of risk management to medical devices*

ISO 80369-1:2010, *Small-bore connectors for liquids and gases in healthcare applications – Part 1: General requirements*

EN 13544-2:2002, *Respiratory therapy equipment – Part 2: Tubing and connectors*
EN 13544-2:2002:AMD 1:2009

ASTM D638-14, *Standard test method for tensile properties of plastics*

ASTM D790-10, *Standard test methods for flexural properties of unreinforced and reinforced plastics and electrical insulating materials*

3 Terms and definitions

For the purposes of this document, the terms and definitions specified in ISO 80369-1:2010, ISO 14971:2007 and the following apply.

NOTE For convenience, the sources of all defined terms used in this document are given in 0.

3.1
NORMAL USE
operation, including routine inspection and adjustments by any USER, and stand-by, according to the instructions for use

Note 1 to entry: NORMAL USE should not be confused with INTENDED USE. While both include the concept of use as intended by the MANUFACTURER, INTENDED USE focuses on the medical purpose while NORMAL USE incorporates not only the medical purpose, but maintenance transport, etc. as well.

[SOURCE: IEC 60601-1:2005+IEC 60601-1:2005/AMD1:2012, definition 3.71, modified, replaced 'OPERATOR' with 'USER'.]

3.2
RATED
<value> term referring to a value assigned by the MANUFACTURER for a specified operating condition

[SOURCE: IEC 60601-1:2005, definition 3.97]

---

2 Both the current and previous versions of this standard are normatively referenced.

3 Both the current and previous versions of this standard are normatively referenced.