TECHNICAL REPORT

ISO/TR 16476

First edition
2016-06-01

Reference materials — Establishing and expressing metrological traceability of quantity values assigned to reference materials

Matériaux de référence — Etablissement et expression de la traçabilité métrologique de valeurs assignées à des matériaux de référence

Reference number
ISO/TR 16476:2016(E)

© ISO 2016



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

 © ISO 2016 – All rights reserved

Contents

Page

Foreword ... **iv**

Introduction ... **v**

1 Scope ... **1**

2 The VIM definition of metrological traceability ... **1**

3 Challenges arising from the definition of metrological traceability ... **2**
3.1 Conventions ... 2
3.2 (C)RM as the carrier of traceable values ... 3
3.3 Implicit traceability to the unit of the measurement scale ... 4
3.4 Traceability networks ... 5
3.5 Properties expressed in units of measurement scales other than the SI ... 5
3.6 Properties other than quantitative ... 6
3.7 Summary of an ISO/REMCO event on metrological traceability ... 6

4 Approaches to metrological traceability of (C)RM ... **7**
4.1 General ... 7
4.2 Approach A ... 7
4.3 Approach B ... 8

5 Establishing traceability of (C)RM property values (Approach B) ... **9**
5.1 Principles ... 9
5.2 Traceability pathways ... 10
5.3 Steps in establishing traceability ... 10
5.3.1 General ... 10
5.3.2 Combining results ... 11
5.4 Summary ... 12

6 Reporting traceability ... **12**
6.1 Inquiry ... 12
6.2 Results of the inquiry ... 12
6.3 Requirements ... 13
6.4 Formats ... 14
6.5 Further recommendations ... 16

Annex A (informative) **Worked-out example** ... **17**

Annex B (informative) **Catalogue of analytes and measurement areas covered by WHO** ... **19**

Annex C (informative) **Example for method-independent, SI traceable values obtained by inter-laboratory comparison** ... **21**

Bibliography ... **22**

© ISO 2016 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

ISO/TR 16476 was prepared by the ISO Committee on Reference Materials (ISO/REMCO).

© ISO 2016 – All rights reserved

Introduction

Reference materials (RM), in particular when certified (CRM), are a major tool for assuring the quality and reliability of results obtained in measurement and testing. CRM property values, in particular used for assessing the trueness of a measurement procedure as implemented in a laboratory, also establish traceability of the measurement result. Which reference the property values assigned to (C)RM should be traceable to, and how this traceability should be established, demonstrated, and reported on certificates is, therefore, a question of primary importance, mainly for RM producers. However, users of (C)RMs should also know what the endpoint of their traceability chain is, in particular for all purposes of cross-border acceptance of measurement results.

It was therefore considered necessary to conduct a study into existing principles for, and requirements to, the traceability of (C)RM, in particular with a specific view to the current definition of metrological traceability given by the Vocabulary of International Metrology (VIM), edition 3, 2007.

© ISO 2016 – All rights reserved

Reference materials — Establishing and expressing metrological traceability of quantity values assigned to reference materials

1 Scope

This Technical Report investigates, discusses, and specifies further, the general principles of establishing traceability of measurement results laid down in the Joint BIPM, OIML, ILAC and ISO Declaration on Metrological Traceability [1], in particular for values assigned to (certified) reference materials. The document covers the following topics:

a) a study into existing principles for, and requirements to, the traceability of the value assigned to the property of a (C)RM, with a specific view to the current definition of metrological traceability given by the 2007 edition of the VIM (published also as JCGM 200:2008[2] and ISO/IEC Guide 99:2007[21]);

b) the development of a sensible, widely applicable approach to the understanding of the traceability of a value assigned to (C)RM property;

c) recommendations on how traceability should be established, demonstrated, and reported on certificates and other documents accompanying (C)RM.

The developed approach is exemplified for measurement procedures not covered earlier by other guidance documents on the topic.

2 The VIM definition of metrological traceability

The recent edition of the VIM[2],[21] defines *metrological traceability* (term 2.41) as

property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty	
NOTE 1	For this definition, a 'reference' can be a definition of a measurement unit through its practical realization, or a measurement procedure including the measurement unit for a non-ordinal quantity, or a measurement standard.
NOTE 2	Metrological traceability requires an established calibration hierarchy.
NOTE 3	Specification of the reference must include the time at which this reference was used in establishing the calibration hierarchy, along with any other relevant metrological information about the reference, such as when the first calibration in the calibration hierarchy was performed.
NOTE 4	For measurements with more than one input quantity in the measurement model, each of the input quantity values should itself be metrologically traceable and the calibration hierarchy involved may form a branched structure or a network. The effort involved in establishing metrological traceability for each input quantity value should be commensurate with its relative contribution to the measurement result.
NOTE 5	Metrological traceability of a measurement result does not ensure that the measurement uncertainty is adequate for a given purpose or that there is an absence of mistakes.

© ISO 2016 – All rights reserved