TECHNICAL REPORT

ISO/TR 19601

First edition
2017-07

Nanotechnologies — Aerosol generation for air exposure studies of nano-objects and their aggregates and agglomerates (NOAA)

Nanotechnologies — Génération d'aérosols pour réaliser des études d'exposition à l'air des nano-objets et de leurs aggrégats et agglomérats (NOAA)

Reference number
ISO/TR 19601:2017(E)

© ISO 2017



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword ..... **v**
Introduction ..... **vi**
1 Scope **1**
2 Normative references **1**
3 Terms and definitions **1**
4 Abbreviated terms **7**
5 Study design considerations **8**
5.1 General 8
5.2 Workplace exposure scenario 8
5.3 Existing inhalation toxicity testing guidelines 9
5.4 Globally harmonized system of classification and labelling of chemicals (GHS) 9
6 Considerations in selection of proper generators **9**
6.1 Basic scheme 9
6.1.1 Flow chart 9
6.1.2 Selection of study 11
6.1.3 Characterization of physicochemical properties of nanomaterials 11
6.1.4 Exposure information on possible use or handling and manufacturing 11
6.1.5 Exposure characteristics 11
6.1.6 Types of inhalation exposure methods 12
6.1.7 Particle characterization method 12
7 NOAA aerosol generators **12**
7.1 General 12
7.2 Dry dissemination 15
7.2.1 Wright dust feeder (see Figure 2) 15
7.2.2 Brush type aerosol generator (see Figure 3) 16
7.2.3 Small scale powder disperser (SSPD; see Figure 4) 17
7.2.4 Fluidized bed aerosol generator (FBG; see Figure 5) 18
7.2.5 Acoustic dry aerosol generator elutriator (ADAGE; see Figure 6) 19
7.2.6 Vilnius aerosol generator (VAG; see Figure 7) 20
7.2.7 Rotating drum generator (see Figure 8) 21
7.3 Wet dissemination 23
7.3.1 Atomizer/nebulizer (see Figure 9 to Figure 11) 23
7.3.2 Electro-static assist axial atomizer (see Figure 12) 26
7.4 Phase change 27
7.4.1 Evaporation/condensation generator (see Figure 13) 27
7.4.2 Spark generator 28
7.4.3 Condensation nano-aerosols 29
7.5 Chemical reaction 30
7.5.1 Combustion 30
7.6 Liquid phase filtration/dispersion — Critical point drying (tertiary butyl alcohol sublimation) and direct injection system for whole-body inhalation studies 32
7.6.1 Principle of operation 32
7.6.2 Advantages 32
7.6.3 Limitations 32
8 Experimental integration **36**
8.1 General 36
8.2 Exposure characterization 37
8.3 Particle properties 37
8.4 Considerations for ***in vivo*** exposure systems 37
9 Considerations in use of nano-aerosol generator for *in vitro* study **38**

© ISO 2017 – All rights reserved

Annex A (informative) **NOAA generators, their particle size distribution and measurement methods** **39**

Annex B (informative) **Aerosol dilution system** **44**

Bibliography **47**

 © ISO 2017 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 229, *Nanotechnologies*.

© ISO 2017 – All rights reserved

Introduction

Inhalation is a primary route of exposure to aerosolized nano-objects and their aggregates and agglomerates (NOAA). The NOAAs include nano-objects with one, two or three external dimensions in the nanoscale from approximately 1 nm to 100 nm, which might be spheres, fibres, tubes and others as primary structures. NOAAs can consist of individual primary structures in the nanoscale and aggregated or agglomerated structures, including those sizes larger than 100 nm. To evaluate the inhalation toxicity of NOAA, it is important to consider certain parameters that make the toxicity testing relevant to human exposure. The three critical aspects to consider when designing and conducting nanomaterial inhalation toxicity study are

a) uniform and reproducible nano-object aerosol generation that is relevant to realistic exposures,

b) thorough characterization of nanomaterials throughout the duration of testing including starting and generated materials, and

c) use of occupational exposure limits (OEL) and reference concentrations (RfC) (as derived from existing studies and/or real-time exposure monitoring data) for dosimetry.

Therefore, to conduct *in vitro* and *in vivo* NOAA, it is important to choose an appropriate NOAA aerosol generator and use online and off-line techniques for nano-object characterization.

Aerosol generation techniques are well established and have been used in laboratory studies, inhalation therapy and industry for many years. A number of aerosol generation techniques are routinely used for other materials that can be adapted for nano-object inhalation toxicity studies. In principle, aerosol generation involves application of some form of energy to the material to reduce its size or to form small particles that are dispersed in a gas stream.

This document provides the status of nano-object aerosol generators. This document further discusses the advantages and limitations of the respective nano-object generators, which can aid in choosing the appropriate generator when conducting the nano-object inhalation toxicity study. No matter what generation system is used for toxicity study, the generated atmospheres should be thoroughly characterized in order to allow for comparison to occupational exposure atmospheres so that a valid risk assessment/occupational exposure limit (OEL) can be developed. Therefore, this document will also provide nano-object aerosol size information generated from respective generators along with the proper nano-object characterization methods. This document complements the work of the Organization for Economic Cooperation and Development (OECD) Working Party on Manufactured Nanomaterial (WPMN) and other related framework documents. Recommendations and guidelines to assist investigators in making appropriate choices of an aerosol generator for their target NOAAs to be tested are presented in this document.

© ISO 2017 – All rights reserved

Nanotechnologies — Aerosol generation for air exposure studies of nano-objects and their aggregates and agglomerates (NOAA)

1 Scope

This document describes methods for producing aerosols of nano-objects and their aggregates and agglomerates (NOAA) for *in vivo* and *in vitro* air exposure studies. The purpose of this document is to aid in selecting an appropriate aerosol generator to fulfil a proposed toxicology study design. This document describes characteristics of aerosol generation methods, including their advantages and limitations. This document does not provide guidance for aerosolization of specific nano-objects.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/TS 80004-1, *Nanotechnologies — Vocabulary — Part 1: Core terms*

ISO/TS 80004-2, *Nanotechnologies — Vocabulary — Part 2: Nano-objects*

ISO/TS 80004-4, *Nanotechnologies — Vocabulary — Part 4: Nanostructured materials*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/TS 80004-1, ISO/TS 80004-2, and ISO/TS 80004-4 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— IEC Electropedia: available at http://www.electropedia.org/

— ISO Online browsing platform: available at http://www.iso.org/obp

3.1
aerodynamic diameter
diameter of a spherical particle with a density of 1 000 kg/m^3 that has the same settling velocity as the *particle* (3.29) under consideration

Note 1 to entry: Aerodynamic diameter is related to the inertial properties of aerosol particles and is generally used to describe particles larger than approximately 100 nm.

[SOURCE: ISO/TR 27628:2007, 2.2]

3.2
aerosol
metastable suspension of solid or liquid *particles* (3.29) in a gas

[SOURCE: ISO/TR 27628:2007, 2.3]

© ISO 2017 – All rights reserved

3.3
agglomerate
collection of weakly bound *particles* (3.29) or *aggregates* (3.4) or *mixtures* (3.17) of the two where the resulting external surface area is similar to the sum of the *surface areas* (3.32) of the individual components

Note 1 to entry: The forces holding an agglomerate together are weak forces, for example, van der Waals forces, or simple physical entanglement.

Note 2 to entry: Agglomerates are also termed *secondary particles* (3.31) and the original source particles are termed *primary particles* (3.30).

[SOURCE: ISO/TS 80004-4:2011, 2.8]

3.4
aggregate
particle (3.29) comprising strongly bonded or fused particles where the resulting external surface area may be significantly smaller than the sum of calculated *surface areas* (3.32) of the individual components

Note 1 to entry: The forces holding an aggregate together are strong forces, for example, covalent bonds, or those resulting from sintering or complex physical entanglement.

Note 2 to entry: Aggregates are also termed *secondary particles* (3.31) and the original source particles are termed *primary particles* (3.30).

[SOURCE: ISO/TS 80004-4:2011, 2.7]

3.5
coagulation
formation of larger *particles* (3.29) through the collision and subsequent adhesion of smaller particles

[SOURCE: ISO/TR 27628:2007, 2.6]

3.6
differential electrical mobility classifier
DEMC
classifier that is able to select aerosol particles according to their electrical mobility and pass them to its exit

Note 1 to entry: A DEMC classifies aerosol particles by balancing the electrical force on each particle with its aerodynamic drag force in an electrical field. Classified particles are in a narrow range of electrical mobility determined by the operating conditions and physical dimensions of the DEMC, while they can have sizes due to difference in the number of charges that they have.

[SOURCE: ISO 15900:2009, 2.7]

3.7
differential mobility analysing system
DMAS
system to measure the size distribution of submicrometre aerosol particles consisting of a *DEMC* (3.6), flow meters, a particle detector, interconnecting plumbing, a computer and suitable software

[SOURCE: ISO 15900:2009, 2.8]

3.8
dustiness
propensity of a material to generate airborne dust during its handling

[SOURCE: EN 1540:2011]

© ISO 2017 – All rights reserved

3.9
engineered nanomaterial
nanomaterial (3.21) that is rationally designed manufactured

[SOURCE: ISO/TS 80004-1:2015, 2.8, modified]

3.10
hazard category
division of criteria within each *hazard class* (3.11) as used in Globally Harmonized System of Classification and Labelling of Chemicals (GHS)

3.11
hazard class
nature of the physical, health or environmental hazard as used in Globally Harmonized System of Classification and Labelling of Chemicals (GHS)

[SOURCE: GHS, 2015]

3.12
geometric mean diameter
GMD
measure of central tendency of particle size distribution using the logarithm of particle diameters

Note 1 to entry: The GMD is normally computed from particle counts and when noted may be based on *surface area* (3.32) or particle volume with appropriate weighting, as:

$$\ln(\mathrm{GMD}) = \frac{\sum_{i=m}^{n} \Delta N_i \ln(d_i)}{N}$$

where

d_i is the midpoint diameter for the size channel, i;

N is the total concentration;

ΔN_i is the concentration within the size channel, i;

m is the first channel;

n is the last channel.

[SOURCE: ISO 10808:2010, 3.5, modified]

3.13
geometric standard deviation
GSD
measure of width or spread of particle sizes, computed for the *DMAS* (3.7) by

$$\ln(\mathrm{GSD}) = \sqrt{\frac{\sum_{i=m}^{n} N_i \left[\ln d_i - \ln(\mathrm{GMD})\right]^2}{N-1}}$$

[SOURCE: ISO 10808:2010, 3.6]

3.14
count median diameter
CMD
diameter equal to *GMD* (3.12) for particle counts assuming a logarithmic normal distribution

[SOURCE: ISO 10808:2010, 3.7, modified]

© ISO 2017 – All rights reserved

3.15
mass median aerodynamic diameter
MMAD
calculated *aerodynamic diameter* (3.1) which divides the *particles* (3.29) of an *aerosol* (3.2) in half based on mass of the particles

Note 1 to entry: 50 % of the particles by mass will be larger than the median diameter and 50 % of the particles will be smaller than the median.

[SOURCE: EPA IRIS Glossary]

3.16
manufactured nanomaterial
nanomaterial (3.21) intentionally produced to have specific properties or composition

[SOURCE: ISO/TS 80004-1:2015, 2.9, modified]

3.17
mixture
solution composed of two or more *substances* (3.33) in which they do not react

Note 1 to entry: A solution is also a mixture.

[SOURCE: GHS, 2015]

3.18
mobility
<aerosols> propensity for an aerosol particle to move in response to an external influence, such as an electrostatic field, thermal field or by diffusion

[SOURCE: ISO/TR 27628:2007, 2.9]

3.19
nano-aerosol
fluid nanodispersion with gaseous matrix and at least one or more liquid or solid nanophase (including *nano-objects* (3.22))

[SOURCE: ISO/TS 80004-4:2015, 3.5.4]

3.20
nanofibre
nano-object (3.22) with two similar external dimensions in the *nanoscale* (3.25) and the third dimension significantly larger

Note 1 to entry: The largest external dimension is not necessarily in the nanoscale.

Note 2 to entry: The terms nanofibril and nanofilament can also be used.

[SOURCE: ISO/TS 80004-2:2015, 4.5, modified]

3.21
nanomaterial
material with any external dimension in the *nanoscale* (3.25) or having internal or surface structure in the nanoscale

Note 1 to entry: Generic term covering both *nano-object* (3.22) and *nanostructured material* (3.27).

EXAMPLE Nanocrystalline materials, nanoparticle powder, materials with nanoscale precipitates, nanoscale films, nano-porous material, nanoscale emulsions and materials with nanoscale textures on the surface. End products containing nanomaterials (e.g. tires, electronic equipment, coated DVDs) are not themselves nanomaterials.

[SOURCE: ISO/TS 80004-1:2015, 2.4, modified]

© ISO 2017 – All rights reserved

3.22
nano-object
discrete piece of material with one, two or three external dimensions in the *nanoscale* (3.25)

Note 1 to entry: The second and third external dimensions are orthogonal to the first dimension and to each other.

[SOURCE: ISO/TS 80004-1:2015, 2.5]

3.23
nanoparticle
nano-object (3.22) with all external dimensions in the *nanoscale* (3.25) where the lengths of the longest and the shortest axes of the nano-object do not differ significantly

Note 1 to entry: If the dimensions differ significantly (typically by more than 3 times), terms such as *nanofibre* (3.20) or *nanoplate* (3.24) may be preferred to the term nanoparticle.

[SOURCE: ISO/TS 80004-2:2015, 4.4]

3.24
nanoplate
nano-object (3.22) with one external dimension in the *nanoscale* (3.25) and the other two external dimensions significantly larger

Note 1 to entry: The larger external dimensions are not necessarily in the nanoscale.

[SOURCE: ISO/TS 80004-2, 4.6, modified]

3.25
nanoscale
size range from approximately 1 nm to 100 nm

Note 1 to entry: Properties that are not extrapolations from a larger size will typically, but not exclusively, be exhibited in this size range. For such properties, the size limits are considered approximate.

Note 2 to entry: The lower limit in this definition (approximately 1 nm) is introduced to avoid single and small groups of atoms from being designated as *nano-objects* (3.22) or elements of *nanostructures* (3.26), which might be implied by the absence of a lower limit.

[SOURCE: ISO/TS 80004-1:2015, 2.1, modified]

3.26
nanostructure
interrelation of the constituent parts of a material in which one or more of those constituent parts belong to the *nanoscale* (3.25)

[SOURCE: ISO/TS 80004-1:2015, 2.6, modified]

3.27
nanostructured material
material having internal or surface structure in the *nanoscale* (3.25)

Note 1 to entry: If external dimension(s) are in the nanoscale, the term *nano-object* (3.22) is recommended.

[SOURCE: ISO/TS 80004-1:2015, 2.7, modified]

3.28
nanotube
hollow nanofibre

[SOURCE: ISO/TS 80004-2:2015, 4.8]

© ISO 2017 – All rights reserved

3.29
particle
minute piece of matter with defined physical boundaries

Note 1 to entry: A physical boundary can also be described as an interface.

Note 2 to entry: A particle can move as a unit.

Note 3 to entry: This general definition applies to particle *nano-objects* (3.22).

[SOURCE: ISO/TS 26824:2013, 1.1]

3.30
primary particle
original source particle of *agglomerates* (3.3) or *aggregates* (3.4) or *mixtures* (3.17) of the two

Note 1 to entry: Constituent particles of agglomerates or aggregates at a certain actual state may be primary particles, but often the constituents are aggregates.

Note 2 to entry: Agglomerates and aggregates are also termed *secondary particles* (3.31).

[SOURCE: ISO 26824:2013, 1.4]

3.31
secondary particle
particle (3.29) formed through chemical reactions in the gas phase (gas to particle conversion)

[SOURCE: ISO/TR 27628:2007, 2.17]

3.32
surface area
area of external surface plus the internal surface of its accessible macro- and mesopore

Note 1 to entry: Includes mass-specific surface area or volume-specific surface area.

[SOURCE: ISO/TR 13014:2012, 2.28]

3.33
substance
chemical elements and their compounds in the natural state or obtained by any production process, including any additive necessary to preserve the stability of the product and any impurities deriving from the process used, but excluding any solvent which may be separated without affecting the stability of the substance or changing its composition

[SOURCE: GHS, 2015]

3.34
ultrafine particle
particle (3.29) with a nominal diameter (such as geometric, aerodynamic, *mobility* (3.18), projected-area or otherwise) of 100 nm or less

Note 1 to entry: The term is often used in the context of particles produced as a by-product of a process (incidental particles), such as welding fume and combustion fume.

[SOURCE: ISO/TR 27628:2007, 2.21]

3.35
reference concentration
benchmark estimates of the quantitative dose–response assessment of chronic non-cancer toxicity for individual inhaled chemicals

[SOURCE: EPA, 1994]

© ISO 2017 – All rights reserved