TECHNICAL REPORT

ISO/TR 230-11

First edition
2018-04

Test code for machine tools —

Part 11: Measuring instruments suitable for machine tool geometry tests

Code d'essai des machines-outils —

Partie 11: Instruments de mesure compatibles avec les essais de géométrie des machines-outils

Reference number
ISO/TR 230-11:2018(E)

© ISO 2018



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

 © ISO 2018 – All rights reserved

Contents

Page

Foreword ..... **v**

Introduction ..... **vi**

1 Scope **1**

2 Normative references **1**

3 Terms and definitions **1**

4 Preliminary remarks **5**
4.1 Measuring units 5
4.2 Uncertainty of measuring instrument 5
4.2.1 General 5
4.2.2 Environment factors 7
4.2.3 Measuring equipment factors 9
4.2.4 Setup and procedure factors 11
4.2.5 Software and calculation factors 17
4.3 Measuring equipment calibration (ISO 10012) 17
4.3.1 General 17
4.3.2 Manufacturer and supplier of measuring instruments 18
4.3.3 User of measuring instruments 18
4.3.4 Measurement uncertainty 18
4.4 Comparison of measurement results by instruments using different measurement principles 18

5 Description of measuring instruments **19**

6 Mechanical artefacts for general use **20**
6.1 General 20
6.2 Reference straightedge 23
6.3 Test mandrels with taper shanks 27
6.4 Mandrels between centres 32
6.5 Reference squares 34
6.6 Reference cube 36
6.7 Surface plates 37
6.8 Reference sphere 38
6.9 1D ball array 39
6.10 2D ball array 42
6.11 Step gauge 43
6.12 Gauge block 44

7 Length and displacement measuring instruments **45**
7.1 General 45
7.2 Instruments for large and medium range linear displacements 46
7.2.1 Laser interferometer 46
7.2.2 Reference linear scale 49
7.2.3 Standard scale 50
7.3 Instruments for short range linear displacements 51
7.3.1 Contact-type sensors 51
7.3.2 Non-contact sensors 55
7.3.3 Contact probing system 60

8 Straightness measuring devices **63**
8.1 General 63
8.2 Taut wire with optical reading device 64
8.3 Laser interferometer with straightness optics 66
8.4 Alignment telescope 67
8.5 Two planes laser scanning device 70
8.6 Alignment laser 71

© ISO 2018 – All rights reserved

9 Squareness measuring devices **72**
9.1 General 72
9.2 Laser interferometer with squareness and straightness optics 73
9.3 Index table with straightedge 74
9.4 Three planes laser scanning device 76

10 Flatness measuring devices **77**
10.1 General 77
10.2 Sweep optical square 78
10.3 Laser interferometer with angular optics 80
10.4 Single plane laser scanning device 81

11 Angle measuring instruments **82**
11.1 General 82
11.2 Level 83
11.2.1 Precision level 83
11.2.2 Electronic level 84
11.2.3 Inclinometer 85
11.3 Autocollimators 86
11.3.1 Autocollimator 86
11.3.2 Photo-electric autocollimators 87
11.3.3 Laser autocollimators 88
11.4 Laser interferometer with angular optics 89
11.5 Reference indexing table with optical angle reading device 90
11.6 Optical polygon with optical reading device 91
11.7 Laser assisted index device 92
11.8 Reference angle encoders 93

12 Special purpose instruments **94**
12.1 General 94
12.2 Single dimensional position reading device 94
12.2.1 Telescoping ball bar 94
12.2.2 Long range telescoping ball bar 96
12.2.3 Fibre interferometer 97
12.3 Multi-dimensional position reading device 98
12.3.1 Two-dimensional digital scale 98
12.3.2 Laser tracing system 99
12.3.3 3D-probe for spheres (contact type) 100
12.3.4 3D-probe head, non-contact type 101
12.4 Other position reading device 103
12.4.1 Spindle error test equipment 103
12.4.2 Swivelling angle measuring device 104

13 Special application examples of the instruments **105**
13.1 Spindle rotation accuracy measurement 105
13.2 Thermal displacement measurements 105
13.3 Tool position set 105
13.4 Workpiece positioning 105

Annex A (informative) **Checking devices for instruments in the workshop** **106**

Annex B (informative) **Existing ISO and national standards for measuring equipment (2015)** **112**

Annex C (informative) **Special caution for supporting system of measuring devices** **113**

Annex D (informative) **Reference table of uncertainty contributor** **115**

Bibliography **126**

 © ISO 2018 – All rights reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 39, *Machine tools,* Subcommittee SC 2, *Test conditions for metal cutting machine tools*.

A list of all parts in the ISO 230 series can be found on the ISO website.

© ISO 2018 – All rights reserved

Introduction

The purpose of this document is to provide information for instruments and equipment for testing machine tools as specified in the ISO 230 series (except ISO 230-5 and ISO/TR 230-8), and in machine-specific standards of ISO/TC 39/SC 2, test conditions for metal cutting machine tools.

The main parts of this document have been transferred from ISO 230-1:1996, Annex A, which is no longer part of ISO 230-1. Newly developed measuring instruments, like special purpose measuring instruments in Clause 12, have been added to this document as well as special application examples in Clause 13.

The concept of measuring uncertainty has been implemented. Uncertainty contributors for measuring instruments and measuring procedures are listed in Annex D to improve reliability of test results. In addition, Annex A addresses checking devices for instruments applied in the workshop and Annex C addresses influences of supporting systems.

Additional information for existing ISO and national standards for measuring equipment is included in Annex B.

This document and ISO 230-1:2012 together cover the entire content of ISO 230-1:1996, with updated instruments and concepts.

© ISO 2018 – All rights reserved

Test code for machine tools —

Part 11: Measuring instruments suitable for machine tool geometry tests

1 Scope

The aim of this document is to document the characteristics of precision measuring instruments for testing the geometric accuracy of machine tools, operating either under no-load or under quasi-static conditions.

Where necessary, reference is made to the appropriate International Standards.

The measuring instruments for operational testing of machine tools [vibrations (ISO/TR 230-8), noise (ISO 230-5), stick-slip motion of components, etc.] as well as instruments for checking of other characteristics of machine tools (speeds, feeds, temperature) are not covered in this document. The measuring instruments for checking of workpiece geometry (size, form, etc.) are not covered by this document either.

This document has list style construction for ease of search and identification of each instrument's characteristics.

Sources of uncertainty of instruments and measurements are described in this document for more accurate measurement procedures.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at https://www.iso.org/obp

— IEC Electropedia: available at https://www.electropedia.org/

3.1
calibration
operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties, and in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

Note 1 to entry: A calibration may be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it may consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system, often mistakenly called "self-calibration", nor with verification of calibration.

© ISO 2018 – All rights reserved

Note 3 to entry: Often, the first step alone in the above definition is perceived as being calibration.

[SOURCE: JCGM 200:2012, 2.39]

3.2
measuring range
set of values of measurands for which the error of a measuring instrument is intended to lie within specified limits

Note 1 to entry: Error is determined in relation to a conventional true value.

[SOURCE: ISO 14978:2006, 3.36]

3.3
accuracy
closeness of agreement between a measured quantity value and a true quantity value of a measurand

Note 1 to entry: The concept "measurement accuracy" is not a quantity and is not given a numerical quantity value. A measurement is said to be more accurate when it offers a smaller measurement error.

Note 2 to entry: The term "measurement accuracy" should not be used for measurement trueness and the term "measurement precision" should not be used for 'measurement accuracy', which, however, is related to both these concepts.

Note 3 to entry: "Measurement accuracy" is sometimes understood as closeness of agreement between measured quantity values that are being attributed to the measurand.

[SOURCE: JCGM 200:2012, 2.13]

3.4
linearity
degree of insignificance on deviation from the linear relation between the input signal and the output signal[20]

3.5
repeatability
measuring precision under a set of repeatability conditions of measurement

Note 1 to entry: These conditions include

— reduction to a minimum of the variations due to the observer,

— the same measurement procedure,

— the same observer,

— the same measuring equipment, used under the same conditions,

— the same location, and

— repetition over a short period of time.

Note 2 to entry: Repeatability can be expressed quantitatively in terms of the dispersion characteristics of the indications.

3.6
frequency response
a state where the amplitude ratio of an output signal to the input signal and the phase difference between the two are varied as a function of sinusoidal input signal frequency[20]

3.7
measuring force
<load> force applied by the stylus of an indicator or recorder to the feature being measured

© ISO 2018 – All rights reserved

3.8
operating environment
atmosphere or environment in which the object is placed during testing[20]

3.9
stability
property of a measuring instrument, whereby its metrological properties remain constant in time

Note 1 to entry: Stability may be quantified in several ways.

EXAMPLE 1 In terms of the duration of a time interval over which a metrological property changes less than a stated amount.

EXAMPLE 2 In terms of the change of a property over a stated time interval.

[SOURCE: JCGM 200:2012, 4.19]

3.10
correction
compensation for an estimated systematic effect

Note 1 to entry: See ISO/IEC Guide 98-3:2008, 3.2.3, for an explanation of "systematic effect".

Note 2 to entry: The compensation value(s) can take different forms, such as a constant addition or multiplication, or multiple values obtained from a table.

[SOURCE: JCGM 200:2012, 2.53]

3.11
measuring instrument
device used for making measurements, alone or in conjunction with one or more supplementary devices

Note 1 to entry: A measuring instrument that can be used alone is a measuring system.

[SOURCE: JCGM 200:2012, 3.1]

3.12
measuring transducer
device, used in measurement, that provides an output quantity having a specified relation to the input quantity

[SOURCE: JCGM 200:2012, 3.7]

3.13
measuring system
set of one or more measuring instruments and often other devices, including any reagent and supply, assembled and adopted to give information used to generate measured quantity values within specified intervals for quantities of specified kinds

[SOURCE: JCGM 200:2012, 3.2]

3.14
sensor
element of a measuring system that is directly affected by the phenomenon, body or substance carrying the quantity to be measured

EXAMPLE Sensing coil of a platinum resistance thermometer, rotor of a turbine flow meter, Bourdon tube of a pressure gauge, float of a level-measuring instrument, photocell of a spectrometer, thermotropic liquid crystal which changes colour as a function of temperature.

Note 1 to entry: In some fields, the term "detector" is used for this concept.

[SOURCE: JCGM 200:2012, 3.8]

© ISO 2018 – All rights reserved

3.15
detector
device or substance that indicates the presence of a phenomenon, body, or substance when a threshold value of an associated quantity is exceeded

EXAMPLE Halogen leak detector, litmus paper.

Note 1 to entry: In some fields, the term "detector" is used for the concept of sensor.

[SOURCE: JCGM 200:2012, 3.9]

3.16
sensitivity
quotient of the change in the indication of a measuring system and the corresponding change in a value of a quantity being measured

Note 1 to entry: Sensitivity of a measuring system can depend on the value of the quantity being measured.

Note 2 to entry: The change considered in a value of a quantity being measured needs to be large compared with the resolution.

[SOURCE: JCGM 200:2012, 4.12]

3.17
resolution
smallest change, in a quantity being measured that causes a perceptible change in the corresponding indication

Note 1 to entry: Resolution can depend on, for example, noise (internal or external) or friction. It can also depend on the value of a quantity being measured.

[SOURCE: JCGM 200:2012, 4.14]

3.18
instrumental drift
continuous or incremental change over time in indication, due to changes in metrological properties of a measuring instrument

Note 1 to entry: Instrumental drift is related neither to a change in a quantity being measured nor to a change of any recognized influence quantity.

[SOURCE: JCGM 200:2012, 4.21]

3.19
optical measuring instrument
instruments, measuring physical, geometrical or material properties based on optical principles, such as photometry, interferometry, geometrical optics, holography, or refractometry[14].

EXAMPLE one-coordinate and multi-coordinate measuring machines, surface-measuring instruments, numerical measuring instruments for machine control, autocollimators, telescopes, contour-measuring instruments.

3.20
maximum permissible error
MPE
<for a metrological characteristic> extreme value of an error of a metrological characteristic permitted by specifications, regulations, etc. for a given piece of measuring equipment

[SOURCE: ISO 14978:2006, 3.21, modified — The domain has been added and "errors" has been changed to "error".]

© ISO 2018 – All rights reserved

3.21
measuring precision
closeness of agreement between indications or measured quantity values obtained by replicate measurements on the same or similar objects under specified conditions

Note 1 to entry: Measurement precision is usually expressed numerically by measures of imprecision, such as standard deviation, variance, or coefficient of variation under the specified conditions of measurement.

Note 2 to entry: The specified condition can be, for example, repeatability conditions of measurement, intermediate precision condition of measurement, or reproducibility condition of measurement (see ISO 5725-1. 1994).

Note 3 to entry: Measurement precision is used to define measurement repeatability, intermediate measurement precision, and measured reproducibility.

Note 4 to entry: Sometimes "measurement precision" is erroneously used to mean measurement accuracy.

[SOURCE: JCGM 200:2012, 2.15]

4 Preliminary remarks

4.1 Measuring units

The units for the following features are:

— displacement, distance and linear deviations: mm or µm;

— angles: degrees or a ratio;

— angular deviation: µm/m or ″ (arc seconds); and

— linear compliance: µm/N.

4.2 Uncertainty of measuring instrument

4.2.1 General

Uncertainty of measuring instrument is a component of combined measurement uncertainty (JCGM 200:2012, 4.24). The instrument uncertainty should be small enough to assess the system performance. The uncertainty of the measurement including the instrument uncertainty should be considered according to ISO 14253-1 ("decision rules") when it is used to check system performance against specifications. However, if the measurement uncertainty is less than 10 % of the specification limit, it is common industrial practice to decide on the conformance or non-conformance based on the indicated measurement value.

Measuring equipment should not be used until it has been allowed to stabilize at the ambient temperature, and stability maintained during the test procedure.

Care should be taken to prevent disturbance to the equipment caused by vibrations, magnetic fields, electrical interference, etc.

General uncertainty contributors are indicated in Table 1 (see also ISO 14253-2). Annex D describes relationship between the instruments and uncertainty contributors. Simple explanation of uncertainty contributors related to the measuring instruments and set-up procedures are described in the following clauses.

© ISO 2018 – All rights reserved